

Certificate of Calibration

Model:		Brand:	
Wrench Capacity:	28-210 N m	Serial number:	001-1121222121840
Date of Caliraton:	2022/12/8	Inspector	Yefeng Sun

	Minimum:	60%	100%
Target torque:	42	126	210
Measured value 1:	42.48	124.81	207.60
Measured value 2:	41.66	126.99	209.61
Measured value 3:	41.75	124.54	207.40
Measured value 4:	41.84	125.87	211.54
Measured value 5:	41.77	126.50	207.60

Testing Standard No.: DIN-ISO-6789

Testing is in compliance with international Standards with test equipment and reference standards calibrated by a laboratory traceable to international standards.

Tester Model: PB6036

Serial No.: 108-2005041

The uncertainty of the test equipment at 20 °C is $\leq \pm 1\%$

