

EC DECLARATION OF CONFORMITY

Name and address of the firm	Shenzhen Jumper Medical Equipment Co., Ltd. D Building, No. 71, Xintian Road, Fuyong Street, Baoan, Shenzhen, Guangdong, China Post Code: 518103
We declare under our sole responsibility that	
the medical device	TENS Therapy Device JPD-ES200, JPD-ES210, JPD-ES220
Name, type or model, batch or serial number, possibly sources and number of items	
of class	Class IIa, Rule 9
According to the classification principle in Annex VIII of Regulation (EU) 2017/745	
SRN	CN-MF-000014343
Basic UDI-DI	69517405ES02R3

meets all the provisions of the Regulation (EU) 2017/745 which apply to it.

Applied harmonised standards, national standards or other normative documents	EN ISO 15223-1:2021 / EN 20417:2021 / EN ISO 14971:2019+A11:2021
	EN ISO 10993-1:2020 / EN ISO 10993-5:2009 / EN ISO 10993-10:2013
	EN 60601-1:2006+A1:2013+A12:2014
	EN 60601-1-2:2015
	EN 60601-2-10:2012 +A1:2016
	EN 60601-1-11:2015
	EN 60601-1-6:2010+A1:2015
	EN 62304:2006+A1:2015
	EN ISO 13485:2016

Conformity assessment procedure		This declaration of conformity is based on the Regulation (EU) 2017/745, Annex IX Section 4, Chapters I and III.	
Notified Body (if consulted)		SGS FIMKO OY NB 0598	
European Representative		MedPath GmbH Mies-van-der-Rohe-Strasse 8 80807 Munich, Germany	
CE certificate No.		FI24/1008005	
2024-03-04	SHENZHEN		

Manufacturer: Shenzhen Jumper Medical Equipment Co., Ltd.

Name and function
Tiffany Zhang, Regulatory Affairs Manager