

Certificate of Calibration

Model:		Brand:	
Wrench Capacity:	3/8 19-110 N m	Serial number:	001-1121223020001
Date of Caliraton:	2023/2/8	Inspector	Yefeng Sun

	Minimum:	60%	100%
Target torque:	19	61	110
Measured value 1:	19.02	60.79	109.10
Measured value 2:	19.06	60.15	110.80
Measured value 3:	19.03	60.96	110.42
Measured value 4:	19.16	61.81	108.67
Measured value 5:	19.27	61.72	109.69

Testing Standard No.: DIN-ISO-6789

Testing is in compliance with international Standards with test equipment and reference standards calibrated by a laboratory traceable to international standards.

Tester Model: PB6036

Serial No.: 108-2005041

The uncertainty of the test equipment at 2σ °C is <±1%

