

Certificate of Calibration

Model:		Brand:	
Wrench Capacity:	5-25 N m	Serial number:	001-1121222121839
Date of Caliraton:	2022/12/8	Inspector	Yefeng Sun

	Minimum:	60%	100%
Target torque:	5	15	25
Measured value 1:	5.05	14.83	25.30
Measured value 2:	4.94	14.81	24.86
Measured value 3:	5.07	15.02	25.11
Measured value 4:	5.01	14.93	25.14
Measured value 5:	5.03	14.88	25.20

Testing Standard No.: DIN-ISO-6789

Testing is in compliance with international Standards with test equipment and reference standards calibrated by a laboratory traceable to international standards.

Tester Model: PB6036

Serial No.: 108-2005041

The uncertainty of the test equipment at 20 °C is $\leq \pm 1\%$

