

# Certificate of Calibration

Model:		Brand:	
Wrench Capacity:	70-350 N m	Serial number:	001-1121222121841
Date of Caliraton:	2022/12/8	Inspector	Yefeng Sun

	Minimum:	60%	100%
Target torque:	70	210	350
Measured value 1:	70.01	207.16	345.93
Measured value 2:	69.39	208.89	350.10
Measured value 3:	69.13	208.24	344.86
Measured value 4:	69.56	207.02	352.75
Measured value 5:	70.51	210.64	351.24

Testing Standard No.: DIN-ISO-6789

Testing is in compliance with international Standards with test equipent and reference standards calibrated by a laboratory traceable to international standards.

Tester Model: PB6036

Serial No.: 108-2005041

The uncertainty of the test equipment at 20 °C is  $\leq \pm 1\%$

