

CalCOFI 9707JD: CTD Oxygen vs Bottle Oxygen

4sec CTD sensor O2 vs bottle O2 (mL/L)

$$y = 1.019x - 0.0034$$
$$R^2 = 0.9985$$

