

# CalCOFI 1301SH: Primary CTD Oxygen vs Bottle Oxygen

4sec ave oxygen sensor vs bottle O2 (ml/l)

$$y = 1.0363x - 0.0383$$

$$R^2 = 0.9994$$

