

CalCOFI 1301SH: Secondary CTD Oxygen vs Bottle Oxygen

4sec ave oxygen sensor vs bottle O2 (ml/l)

$$y = 1.0334x - 0.033$$

$$R^2 = 0.9994$$

