

CalCOFI 1504NH: Secondary CTD Oxygen vs Bottle Oxygen

4sec average sensor O2 vs bottle O2 (ml/L)

$$y = 1.2892x + 0.0202$$
$$R^2 = 0.9988$$

