

20000013222a

CBC 5 control_062017

IKA®

IKA® CBC 5 control




Operating instructions

EN



IKA®-Werke, Germany
Reg. No. 004343

Device setup



Item	Designation
1	Wireless Controller (WiCo)
2	Filling opening lid
3	Front cover
4	Handle
5	Venting grid
6	Mains switch
7	Power socket
8	Cable clip
9	Pump connection IN
10	Pump connection OUT
11	RS 232 interface
12	USB interface
13	External temperature sensor socket
14	Multifunction interface

Contents

	Page
Device setup	2
Declaration of Conformity	4
Note for USA (FCC)	4
Note for Canada (IC)	4
Explication of warning symbols	4
Safety instructions	5
General information	5
Fluids.....	6
Battery pack RB 1 (for WiCo)	6
Correct use	7
Use	7
Range of use (indoor use only).....	7
Wireless remote control.....	7
Unpacking	8
Unpacking	8
Delivery scope	8
Preparations	8
Setting up	8
Connecting the tubings	8
Inserting battery into the WiCo.....	9
Mounting the WiCo to the station	9
Filling and draining	9
Fluid (Standard information for IKA® fluid).....	10
Moving the device.....	11
Charging the Battery Pack RB 1 (rechargeable battery)	11
Changing battery to WiCo	11
WiCo – holder WH 10	11
Operator panel and display	12
station.....	12
WiCo.....	12
Setting the safety temperature	13
Commissioning	13
Useful information	14
Working with WiCo	15
Working screen at the time of delivery.....	15
Explanation of symbols on the working screen.....	15
Menu navigation and structure	16
Menu (Details)	18
Interface and output	23
Maintenance and cleaning	26
Error codes	27
Accessories	28
Technical data	29
Warranty	31
Pump performance curve	31

Declaration of Conformity

We declare under our sole responsibility that this product corresponds to the regulations 2014/35/EU, 2014/30/EU and 2011/65/EU and conforms with the standards or standardized documents: EN 61010-1, EN 61010-2-010, EN 61010-2-051, EN 61326-1, EN 60529 and DIN 12876-1.

Bluetooth® module:

Directive: 2014/53/EU

Standards: EN 300328, EN 301489-17, EN 301489-1, EN 60950-1

Note for USA (FCC)

This equipment complies with Part 15 of the FCC rules. Any changes or modifications not expressly approved by the Manufacturer could void the user's authority to operate the equipment. This device complies with Part 15 of the FCC rules subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) This device must accept all interference received, including interference that may cause undesired operation.

NOTE:

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment.

This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.

Note for Canada (IC)

This device complies with Industry Canada license-exempt RSS standard(s). Operation is subject to the following two conditions:

- (1) This device may not cause interference
- (2) This device must accept any interference, including interference that may cause undesired operation of the device.

This device complies with Health Canada's Safety Code 6/IC RSS-210. The installer of this device should ensure that RF radiation is not emitted in excess of the Health Canada's requirement.

Explication of warning symbols



Indicates an (extremely) hazardous situation, which, if not avoided, will result in death, serious injury.



Indicates a hazardous situation, which, if not avoided, can result in death, serious injury.



Indicates a potentially hazardous situation, which, if not avoided, can result in injury.



Indicates practices which, if not avoided, can result in equipment damage.



Indicates a hazardous situation that cause from a hot surface!

Safety instructions

General information:

- **Read the operating instructions completely before starting up and follow the safety instructions.**
- Keep the operating instructions in a place where it can be accessed by everyone.
- Ensure that only trained staff work with the device.
- Follow the safety instructions, guidelines, occupational health, safety and accident prevention regulations.
- Set up the device in a spacious area on an even, stable, clean, non-slip, dry and fireproof surface.
- Do not operate the device in explosive atmospheres, with hazardous substances.
- Protect the device and accessories from bumping and impacting.
- Check the device and accessories for damage before each use. Do not use damaged components.
- Safe operation is only guaranteed with the accessories described in the "Accessories" section.
- The device must only be operated with the original mains cord.
- The socket for the mains cord must be easily accessible.
- Socket must be earthed (protective ground contact).
- The voltage stated on the type plate must correspond to the mains voltage.
- The device can only be disconnected from the mains supply by pulling out the mains plug or the connector plug.
- Disconnect the mains plug before attaching or changing any accessories.
- Disconnect the mains plug before cleaning, maintenance or moving the device.
- The device must only be opened by trained specialists, even during repair. The device must be unplugged from the power supply before opening. Live parts inside the device may still be live for some time after unplugging from the power supply.



NOTICE

Coverings or parts that can be removed from the device must be put back on the device again to ensure safe operation, for example to keep foreign objects and liquids, etc. from getting into the device.

- The device must only be used as prescribed and as described in the operating instructions. This includes operation by instructed specialist personnel.
- When using critical or hazardous materials in your processes, **IKA®** recommends to use additional appropriate measures to ensure safety in the experiment. For example, users can implement measures that inhibit fire or explosions or comprehensive monitoring equipment.
- Process pathogenic material only in closed vessels under a suitable fume hood. Please contact **IKA®** application support if you have any question.



CAUTION

If the mains switch is not within reach when device is operating, an **EMERGENCY STOP** switch that can be easily accessed must be installed in the work area.

- A laboratory circulator heats/refrigerates and circulates fluid according to specified parameters. This involves hazards due to high and low temperatures, fire and general hazards due to the device of electrical energy. The user safety can not be ensured simply with design requirements on the part of the device. Further hazard sources may arise due to the type of tempering fluid, e.g. by exceeding or undercutting certain temperature thresholds or by the breakage of the container and reaction with the carrier fluid. It is not possible to consider all eventualities. They remain largely subject to the judgment and responsibility of the operator. For this reason, it may become necessary for user to take other precautionary safety measures.
- Insufficient ventilation may result in the formation of explosive mixtures. Only use the device in well ventilated areas.



WARNING

The safety circuit (safe temperature) must be adjusted so that the maximum permissible temperature cannot be exceeded even in the event of a fault. Check the safe temperature circuit on a regular basis (see section "Setting the safety temperature").

- When device is used for external circulation, extra precaution must be taken for hot/cold liquid leakage due to damaged hose:
 - Use suitable hoses for connection.
 - Secure hoses and tubes against slippage and avoid kinks.
 - Check hoses, tubes and bath at regular intervals for possible material fatigue (cracks/leaks).
- Mains cable should not get in contact with hot parts and fluids.



WARNING

Do not start up the device if:

- It is damaged or leaking
- Cable (not only mains cable) is damaged.
- Be careful when refilling a hot bath.



CAUTION

At high operating temperature, the temperature of housing parts, surfaces and tubes can exceed 70 °C.



CAUTION

It is dangerous to touch the heater. The temperature of the heater can be very high.

- After a power failure during operation, the device may start automatically (depending on operating mode).
- Transport the device with care.
- Do not transport or empty the bath while it is still hot/cold. This may result in accidents, especially scalding injuries or frostbite.



Because of the heavy weight of the device, at least two persons are needed for carrying the device.

Fluids:



Beware of the risk of burning due to delay in boiling!



Only use fluids, which fulfill the requirements for safety, health and device compatibility. Be aware of the chemical hazards that may be associated with the bath fluid used. Observe all safety warnings for the fluids.

- Depending on the bath fluid used and the type of operation, toxic or flammable vapors can arise. Ensure suitable extraction.
- Do not use any fluid which may cause dangerous reactions during processing.
- Only use recommended bath fluid. Only use non-acid and non corroding fluid.



Only process and heat up any fluid that has a flash point higher than the adjusted safe temperature limit that has been set. The safe temperature limit must always be set to at least 25 °C lower than the flash point of the fluid used. Examine regularly the function of the safety temperature limiter.



Never operate the device without sufficient heat carrier fluid! Check the fluid level detection at a regular basis (see the chapter "Filling and draining").

- Continuous monitoring of the bath and the filling level of the bath fluid is required, especially at high temperatures.
- To ensure a sufficient fluid circulation, the viscosity of the bath fluid at the lowest operating temperature must not exceed a value of 50 mm²/s.



If water is used at higher temperature, there is heavy loss of fluid due to the evaporation of the steam.

- Untreated tap water is not recommended. It is recommended to use distilled water or high purity water (ion exchangers) and add 0.1 g soda (sodium carbonate Na₂CO₃) /liter, to reduce corrosive properties.



Risk of burning caused by vapor or hot water at the outlet of the cooling coil.



Don't use following fluids:

- Untreated tap water
- Acids or bases
- Solutions with halides: chlorides, fluorides, bromides, iodides or sulfur
- Bleach (Sodium Hypochlorite)
- Solution with chromates or chromium salts
- Glycerine
- Ferrous water.



When changing the bath fluid, remove the remaining water from the complete system (including hoses and external devices). When doing this, also open the stopper and union nuts caps of the pump outputs and inputs and blow compressed air through all the pump outputs and inputs! Beware of the risk of burning due to delay in boiling!

Battery pack RB 1 (for WiCo):



If during operation the battery pack RB 1 (rechargeable battery) becomes fully discharged, the device (Station) will continue to run or is shut down depending on the value settings for "Time Out", "Safe Speed" and "Safe Temperature". If the device is set so that it continues to run when the battery of the WiCo is fully discharged, the only means of switching the station off are the "safe STOP", "ON/OFF" key and the "Mains switch"!



Please note the following safety instructions for the battery pack RB 1 (rechargeable battery):

- Keep the battery pack out of reach of children at all times.
- Store the battery pack in a cool, dry place.
- Never throw the battery pack into a fire. Keep it away from direct sunlight and temperatures above 60 °C. High temperatures will damage the battery pack and render it unusable. Temperatures above 100 °C may cause it to explode.
- Never throw the battery pack into water or expose it to moisture. Water may lead to a short-circuit, causing the battery pack to explode.
- Do not deform or crush the battery pack or damage it in any other way. This can cause battery fluid to leak and/or the battery pack to explode.
- When not in use, keep battery packs away from paperclips, coins, keys, nails, screws or other small metal objects which could cause the contacts to be bridged. Short-circuiting may result in an explosion.

- Explosion of a battery pack may release battery fluid and cause a fire.
- The lithium polymer battery pack must only be used and charged in **IKA®** products designed for use with this battery pack.
- When the battery pack is inserted it should slide in easily and without resistance. Do not force it.
- If the battery pack is removed for an extended period of time, store it in a sealed plastic bag to prevent short-circuiting due to moisture or contact with metal components.
- The operating temperature range of the battery pack is from 0 °C to + 45 °C. Note that the battery pack capacity will be reduced at temperatures below 20 °C.
- Only the rechargeable battery types recommended in the technical data may be used in the device!



Do not charge batteries that have leaked or that are discolored, deformed or damaged in any other way.

Disposal instructions:

- When disposing of the **IKA®** battery pack, please tape over the contacts with adhesive tape to prevent short-circuiting due to moisture or contact with metal components. Short-circuiting may result in an explosion.
- Do not throw used battery packs into your household waste. Dispose of them properly in accordance with statutory regulations.



End users are obliged by law to return all used disposable and rechargeable batteries. Throwing them into the household waste is prohibited. Disposable/rechargeable batteries containing harmful substances are marked with this symbol to indicate that they may not be disposed of as household waste.

- You can return used disposable and rechargeable batteries free of charge to your local authority collection site or to any battery retailer. In doing so you will be complying with statutory regulations and helping to protect the environment.
- Batteries must be disposed of in accordance with local and national regulations.

Correct use

• Use:

Use CBC (Combined refrigerating and heating Bath Circulators) for heating/refrigerating and circulating fluids.

Intended Use: Tabletop device

• Range of use (indoor use only):

- Laboratories - Schools
- Pharmacies - Universities

• Wireless remote control:

Before using the wireless link between the WiCo and the laboratory device, first check whether your region is included in the radio communication approval for the device. If it is not, remote control can also be performed using a USB cable.

This device is suitable for use in all areas except:

- Residential areas
- Areas that are connected directly to a low-voltage supply network that also supplies residential areas.

The safety of the user cannot be guaranteed:

- If the device is operated with accessories that are not supplied or recommended by the **IKA®**.
- If the device is operated improperly or in contrary to the **IKA®** specifications.
- If the device or the printed circuit board are modified by third parties.

Unpacking

• Unpacking:

- Please unpack the device carefully.
- In the case of any damage a detailed report must be sent immediately (post, rail or forwarder).



NOTICE

Transport safety:

Open the filling open lid (2) and remove the protection under the buoyage.

• Delivery scope:

- CBC 5 control station with WiCo
- Mains cables
- Hose olive NW 12 (2 pieces) see Fig. 2
- Screwdriver (use for safety circuit) see Fig. 3
- WiCo holder WH 10
- OS 1.0 power supply (for WiCo) see Fig. 4
- RB 1 battery pack (for WiCo)
- USB 2.0 cable micro A – micro B
- USB 2.0 cable A – micro B
- User guide
- Warranty card.




Fig. 2




Fig. 3




Fig. 4

Preparations

• Setting up:

- Place the unit on an even, stable, clean, nonslip, dry and fireproof surface.
- Keep at least 20 cm of open space on the front and rear side.
- The place for installation should be large enough and provide sufficient air ventilation to ensure the room does not warm up excessively because of the heat from device radiates to the environment.
- Do not set up the device in the immediate vicinity of heat sources and do not expose to sun light.
- Cooling machine, pump motor and electronics produce intrinsic heat that is dissipated via the venting grids (5)! Never cover these venting grids!

Note: After setting up the device, wait at least one hour before starting the operation to avoid the damage to the cooling system.

• Connecting the tubings:

- Unscrew the union nuts and stoppers using a wrench (SW19) from the pump connection IN (9) and OUT (10).




Fig. 5

- Connect the hoses for circulating the external system to the pump connection M 16 x 1 for IN and OUT directly or with the olives.
- Screw the hose olives to the pump connection IN and OUT with union nuts. Slide the hoses (NW 12) onto the olives. The hoses must be secured with suitable clamps.





Fig. 6


Note: Please note the permissible temperature range of hoses. For hot fluids we recommend the IKA® LT 5.20 hoses.

When the external system is not necessary, please seal the pump connection IN and OUT with the existing union nuts and stoppers.

• Inserting battery into the WiCo:




• Mounting the WiCo to the station:



• Filling and draining:


- Before filling the fluid into the bath, open the front cover as indicated in following figure.




- Check and make sure that the drain valve is closed (Rotate clockwise to the stop position, see Fig. 10).

Note: Please note information in section "Commissioning".

- Connect the mains plug and turn on the device with mains switch (6).
- The low level icon () appears on the display of the WiCo.
- Open the filling lid (2) and fill fluid to the bath.



- To drain the fluid from the bath, connect a hose to the drain port and turn the drain valve in counter clockwise direction with a straight screwdriver.




• **Fluid (Standard information for IKA® fluid):**

IKA® Designation	Operating temperature range for open bath application (°C)	Operating temperature range for closed bath applications (°C)	Safety temperature (°C)	Flash point (°C)
CF.EG28.N10.80.8	-10 ... 80	-10 ... 80	90	115
CF.EG39.N20.80.16	-20 ... 80	-20 ... 80	90	115
CF.EG44.N25.80.19	-25 ... 80	-25 ... 80	90	115
CF.EG48.N30.80.22	-30 ... 80	-30 ... 80	90	115
UF.Si.N30.150.10LV	-30 ... 130	-30 ... 150	145 ①	>170
HF.Si.20.200.50	20 ... 200	20 ... 200	255	>280
HF.Si.20.250.50A	20 ... 200	20 ... 250	255	>280
H ₂ O (Water) ②	5 ... 95	5 ... 95	-	-
Customized 1 ③				
Customized 2 ③				

Check the suitability of the fluid according to your application.

Nomenclature for **IKA®** fluid:

CF.EG28.N10.80.8 --



(1) Classification:

HF: Heating Fluid
 CF: Cooling Fluid
 UF: Universal Fluid

(2) Chemical composition:

Si: Silicone oil
 EG: Ethylene Glycol

(3) Temperature range: (Minimum temperature. Maximum temperature)

N: Negative Temperature

(4) Viscosity:

Viscosity at 25 °C for Heating Fluid (HF)
 Viscosity at -20 °C for Cooling Fluid (CF)
 Viscosity at 25 °C for Universal Fluid (UF)

Dynamic viscosity [mPa•s] is a product of kinematic viscosity [mm²/s] and density [kg/m³] of the fluid divided by 1000.

(5) Additional information:

A: Oil Additives
 LV: Low Viscosity

① Note: for open bath application!

② Note: Tap water may be unsuitable for operation because the calcium carbonate content may cause calcification. High purity water (from ion exchangers) and distilled or bi-distilled water are unsuitable for operation due to corrosive properties of these media. High purity water and distillates are suitable as a medium after adding 0.1 g soda (Na₂CO₃, sodium carbonate) per liter of water.

③ Note: The limits can be adjusted according to the fluid used.

- **Moving the device:**

Empty all fluid in the bath before moving device from one place to other place.

The device must be lifted by two persons with the upper handles. For this purpose, take off the front cover (3).

It can also be moved on flat surface by lifting and pushing the front of the device. It is easy to move the device with the help of the wheels.

The angle of inclination should never be more than 10 ° at any direction when move the device!

Note: The device must not be moved during operation. After moving the device, you must wait at least one hour before restarting the unit.




Fig. 12

- **Charging the Battery Pack RB 1 (rechargeable battery):**

The battery pack in the WiCo can be charged by any of the following means:

- On the station
- Via a USB cable at the PC or station
- Via an OS 1.0 power supply unit.

- **Changing battery to WiCo:**




Fig. 13

Note: Please comply with the relevant safety instructions in the "Safety instructions" section for the RB 1 battery pack!

- **WiCo – holder WH 10:**




Fig. 14

Operator panel and display

• station:




Fig. 15

Item	Designation	Function
A	On/Off key:	Switch on/off the station.
B	Adjustable safety circuit:	Adjust the safety temperature limit with delivered screwdriver.
C	LED bar:	Display different status of the circulator with different color.
D	Contacts:	Communicate and charge the WiCo.
E	"safe STOP" key:	Safe stop the working of the station in emergency.
F	Bluetooth® searching key:	Search the WiCo when the Bluetooth® is active.
G	Bluetooth® LED:	Bluetooth® indicator.

• WiCo:



Fig. 16

Item	Designation	Function
L	On/Off key:	Switch the WiCo on and off.
M	"int/ext" key:	Switch between the internal and external temperature display and control.
N	"Timer/Pump" key:	Switch between the timer and pump display.
O	Graph key:	Display time/temperature graph.
P	Rotating/pressing knob:	Set the temperature value.
R	Rotating/pressing knob:	Start/Stop the heating/cooling function. Navigation, selecting and changing the settings in the menu. Set the pump speed value. Start/Stop the pump function.
S	Display:	Screen.
T	"Back" key:	Return to the previous menu level.
U	"Menu" key:	Press it once: main menu is displayed. Press it a second time: back to the working screen.
V	Lock key:	Lock/Unlock the knobs and keys.

Setting the safety temperature

Setting the safety temperature with screwdriver delivered with the device.




Fig. 17

The safety temperature setting will appear on the display.


Factory setting: maximum value
Adjustment range: see 0 to 260 °C.

Note: The safe temperature limit must always be set to at least 25 °C lower than the flash point of the fluid used.

Commissioning

Note: Before commissioning, make sure that the device has not been moved in one hour!

Check whether the voltage given on the type plate corresponds to the available mains voltage.



The socket used must be earthed (fitted with earth contact).

If these conditions have been met, the machine is ready for operation when the mains plug is plugged in.

If these conditions are not met, safe operation is not guaranteed and the machine could be damaged.

Observe the ambient conditions (temperature, humidity, etc.) listed under "Technical Data".

After pressing the mains switch (6) of the station, the screen of the WiCo displays the device designation and the software version after a beep.



Fig. 18

After several seconds, screen display the information of WiCo.

Information	
Internal Max:	95 °C
Internal Min:	5 °C
External Max:	95 °C
External Min:	5 °C
Fluid:	H ₂ O (Water)

Fig. 19

Then the working screen appears and the device is ready for operation.

A	
0.00	
int °C	
- - -	Safe Temp °C
ext °C	
int	5.00
set °C	
	1000
	Pump rpm

Fig. 20

Change the temperature setting with left knob (P).

Change the pump speed setting with the right knob (R).

In standby status, activate the heating/refrigerating function by pressing the left knob (P). The device start heating/refrigerating function. Meanwhile, the pump starts to run at once.

After pressing the mains switch (6) the start-up of the compressor is delayed (about 3 minutes).

In working status, press the right knob (R) to stop operation of the pump. The heating/refrigerating function and the pump stops.

Note: In standby status, press the right knob (R) to start the pump function, the heating/refrigerating function will not be activated.

In working status, press the left knob (P) to stop the heating/refrigerating, the pump keep running.

When the WiCo is switched on without being connected to the station, the green LED bar (C) and the green Bluetooth® LED (G) on station lights up.

The control elements of the WiCo allow to be locked by pressing the lock key (V) about 2 seconds, so no accidental changes during operation are possible (lock symbol  appears in the display).

By pressing lock key (V) about 2 seconds again, the control elements are released (lock symbol  disappears from the display).

Note: In an emergency, the device function can be turned off by pressing the "safe STOP" key (E) at the front of station. The LED bar (C) changed into red color.

A message appears in the display indicates that the station was forced off. To restart, press the mains switch (6) or On/Off key (A) off and on again.

If the Bluetooth® function of the WiCo is activated, the Bluetooth® icon  appears on the screen and the user can search the WiCo by pressing the searching key (F). Then, a beep is heard.

Useful information

The station is controlled via a WiCo. If the WiCo is mounted on the station, data is exchanged between the station and WiCo via the contacts (D). The screen of the WiCo displays the home icon .

The WiCo is equipped with a USB socket (Universal Serial Bus) with which the WiCo can be connected with station, the USB icon  appears on the screen.

If the WiCo is not connected with station via a USB cable, the data exchange between the station and the WiCo via Bluetooth®. In this case, the Bluetooth® icon  is displayed.

Depending on the structure of the building, the WiCo can be operated at a distance up to 15 m from the station, using the Bluetooth® connection.

The WiCo could be either installed on the station or put on a safe place where is accessed easily by the user during operation.

If the WiCo is mounted on the station, the battery is charged through the contacts (D).

The battery could also be charged via the USB port on the WiCo (See "Charging the Battery Pack RB 1" in "Preparations" section).

• **Working screen at the time of delivery:**




Fig. 21

Note: The wireless symbol  appears only when the station is switched on.

• **Explanation of symbols on the working screen:**

The symbols displayed change depending on the status and settings of the WiCo and the station. The screen below shows the most significant symbols on the working screen.




Fig. 22

 Bluetooth®:

This symbol means the station and the WiCo are communicating via Bluetooth®. The symbol no longer appears if no Bluetooth® communication is being performed.

99 Device number:

This figure appears when the "Device Number" is activated in menu option "Display". The figure can be set from 0 to 99. See section "Display" under "Menu (Details)".

 Lock:

This symbol means that the function of the keys and the rotary knobs for controlling the WiCo are disabled by pressing lock key about 2 seconds. The symbol no longer appears if the functions are enabled once again by pressing the lock key about 2 seconds again.

 Temperature Sensor:

This symbol appears when the external temperature sensor is connected.

A Operating Mode:

This symbol indicates the operating mode currently selected (A, B, C, D).

 USB:

This symbol means the WiCo is communicating or the battery is charged via a USB cable. The symbol no longer appears if no USB cable is being used for communicating with the station.

 Home:

This symbol means that the WiCo is connected to the station and is communicating with the station via the contacts.

The symbol no longer appears if the WiCo is removed from the station.

 Battery pack:

This symbol indicates the charging status of the RB 1 battery pack within the WiCo.

The charging symbol appears if the WiCo

- is connected to a PC via a USB cable
- is connected to a station via a USB cable
- is connected to the power supply unit OS 1.0 via a USB cable
- is connected to the station via the charger contacts.

 Heating

This symbol indicates that the heating function is active.

 →  →  indicate active heating process.

 Refrigerating

This symbol indicates that the refrigerating function is active.

 →  indicates active refrigerating process.

 Fluid level

This symbol indicates fluid level.

The red symbol  means the fluid is above the maximum fluid level. Excess fluid should be drained out.

The red symbol  means the fluid is below the minimum fluid level. Please add fluid.

 Pump:

This symbol indicates that the pump is activated.

 Warning:

This symbol indicates that warning is active.

PC PC Control:

This symbol means that either the station or the WiCo is connected to a computer and is being controlled from the computer.

PR Program Control:

This symbol indicates that the WiCo is controlled by a program (see "PROGRAMS").

• Menu navigation and structure:

Menu navigation:



Fig. 23


Control elements
for menu navigation

- ☞ Press the "Menu" key (U).
- ☞ Select the menu by turning the right rotating/pressing knob (R) to the right or left.
- ☞ Open the menu item by pressing the right rotating/pressing knob (R).
- ☞ Turn the rotating/pressing knob (R) to select the desired menu option and edit the values or settings.
- ☞ Press the rotating/pressing knob (R) to get into sub menu items to active/inactive to switch settings or to confirm settings ("OK").
- ☞ Press the "Back" key (T) to a setting or cancel to return to the previous menu.
- ☞ Press the "Menu" key (U) to return directly to the working screen.

Note: When the tempering or pump functions are active, the menu is locked. On the display, the active menu option is highlighted in yellow. The active status of a menu item is marked with a check (✓).

Menu navigation: → Press the "Menu" key (U) and rotate rotating/pressing knob (R)

← Press the "Back" key (T) or the "Menu" key (U)



Note: If you press the "Menu" key (U), the system skips directly back to the working screen.
If you press the "Back" key (T), the system skips back to the previous screen.

Fig. 24

Menu structure:

			Factory Settings
 TEMPERING	Control	Control Mode	Internal (int)
		Control Mode	External (ext)
		Control Parameters	Automatic
			Manual
		Manual	Accurate
			Fast
		Manual	Internal (Kp, Ti, Td, Ts, Prop_Bp, Prop_Bn)
			External (Kp, Ti, Td, Ts, Prop_Bp, Prop_Bn)
			Information (Kp, Ti, Td, Ts, Prop_Bp, Prop_Bn)
			Information (Kp, Ti, Td, Ts, Prop_Bp, Prop_Bn)
		Fluids	CF.EG28.N10.80.8
			CF.EG39.N20.80.16
			CF.EG44.N25.80.19
			CF.EG48.N30.80.22
			UF.Si.N30.150.10LV
			HF.Si.20.200.50
			HF.Si.20.250.50A
			H ₂ O (Water)
			Customized 1
			Customized 2
		Cooling	activated
		Limits	Maximum
			Minimum
		Δ T (int - ext)	100 °C
		Power Output	Heating
			100 %
		Calibration	Internal
			2 Point Calibration
			3 Point Calibration
			External
		Reset Calibration	Internal
			External
		Notification	ON
			Hysteresis
			Sound
			Infinite
 PUMP	Limits	Speed	activated
		Maximum	4600 rpm
		Minimum	1000 rpm
		Pressure	0.7 bar
 TIMER	Set	Display	00:00 [hh:mm]
		A	activated
		B	-
		C	-
 MODE	Display	D	-
		A	-
		B	-
		C	-
 DISPLAY	Main Screen	Pump	activated
		Timer	-
		Device number	-
		1	-
 GRAPH	Automatic	Axis Assignment	Set Temperature
			Actual Internal Temp.
			Actual External Temp.
			Actual External Temp.
		Axis Scaling	X - Axis
			15 min
			30 min
			60 min
			120 min
		Y - Axis	Maximum
			95 °C
			Minimum
			5 °C
 PROGRAMS	Program 1	Program 10	-
		Program 10	-
		Program 10	-
		Program 10	-
 MODULES	M1 - Valve	ON	activated
			Default Status
			Open
			Closed
		Refill	Level 2
			Level 4
			95 °C
			5 °C
		Temperature (Int)	Higher
			Lower
			Int > Set Temperature
			Error
		Alarm	Warning
			Default Status
			Open
			Closed
		Switch	ON
			Default Status
			Open
			Closed
		Temperature (Int)	Higher
			Lower
			Int > Set Temperature
			Ext. Standby
		M3 - Input	ON
			Default Status
			Open
			Closed
 OUTGAS	Speed	Interval	1000 rpm
		Interval	10 s
		Time out	00:30 [mm:ss]
		Set	1000 rpm
 SAFETY	Temperature	Speed	30.00 °C
		Temperature	000
		Safe Temp. Confirmation	-
		Safe Temp. Confirmation	-
 SETTING	Languages	English	activated
		Deutsch	-
			-
		Display	activated
		Background	-
		White	-
		Brightness	100%
		Standard Mode	-
		Battery Mode	40%
		Firmware Update Info	-
		Key Tone	-
		Sound	-
		Factory Settings	-
		Communication	-
		Device name	Circulator WiCo
		Bluetooth	activated
		Tempering	yes
		Pump	yes
		Safety	yes
		Software	yes
		Intro	yes
		Firemware Update Info	yes

• **Menu (Details):**



TEMPERING:

Control:

Control Mode:

Internal (int): The temperature is regulated according to the internal temperature sensor.

Extern (ext): The temperature is regulated according to the external temperature sensor.

Control parameters:

Automatic: The optimal control parameters for PID temperature control are determined automatically. This is the recommended mode.

Selecting "Automatic" allows you to set the dynamics for temperature control.

Accurate: precise tempering without overshooting.

Fast: fast tempering with minimal overshooting.




Fig. 25 (heating curve in "Automatic" mode)

Manual: The control parameters for PID temperature control can be set manually.

"Manual" should only be used in the event of special temperature control requirements.

When "Manual" is selected, the following parameters can be set for "Internal (int)" and "External (ext)" temperature control:

Kp: Proportional coefficient

The proportional coefficient **Kp** is the controller amplification and determines how strongly the control deviation (the difference between the target temperature and actual temperature) directly affects the control variable (on-time of the cooling). **Kp**-values that are too large can lead to the controller overshooting.

Ti: Integral time

The integral time **Ti** (s) is the correction time and determines how strongly the duration of the control deviation affects the control variable. **Ti** compensates for an existing control deviation. A high **Ti** means a smaller and slower effect on the control variable. **Ti**-values that are too small can lead to instability of the controller.

Td: Differential time

The differential time **Td** (s) is the derivative time and determines how strongly the rate of change of the control deviation affects the control variable. **Td** compensates for rapid control deviations. A high **Td** means a smaller and slower effect on the control variable. **Td**-values that are too large can lead to instability of the controller.

Ts: Sampling time

The sampling time **Ts** (s) is the time interval over which the control deviation is determined and the respective control variable (dependent on **Kp**, **Ti** and **Td**) is calculated.

Ts must be adjusted to match the response characteristic (total of all time constants) of the closed loop controlled system, so that the control variable can deliver a uniform and measurable change in the control deviation. **Ts** values that are too small or too large can lead to instability of the controller.

Prop_Bp: Proportional Band Maximum.

Prop_Bn: Proportional Band Minimum.

The Proportional Band is the range below (Prop_Bp) and above (Prop_Bn) the set value in which the control output value is calculated via the difference between the actual and the set value and the PID parameters.

Examples of non-optimal settings:




Fig. 26

Fluids:

Under the option "Fluids", a variety of fluids can be selected.

The selected fluid limits the setting range of the target temperature. See table in the section "Fluid (Standard information for IKA® fluid)".

The maximum and minimum temperature values of the selected fluid can be set within these limitations.

Cooling:

This menu option allows you to activate/deactivate the Cooling function.

Limits:

Under the option "External (ext)", the maximum and minimum temperature for external temperature control can be limited.

Under " ΔT (int - ext)", the maximum difference between the internal and external temperature can be set. The function limits the absolute temperature difference between the external system and internal bath during the heating up and cooling down process. This can protect sensitive device like the glass reactors from thermal shock (e.g. glass break).

Under "Power Output", the maximum heating output can be set as a percentage of nominal heat output.

Temperature sensor:

Calibration:

The internal and the external temperature measurement can be calibrated and adjusted.

You can select 2-point calibration or 3-point calibration for internal and external measurement.

Calibration proceeding (example: 2-point calibration):



Fig. 27

Reset Calibration:

By "Reset Calibration", the calibration value for the internal or the external temperature sensor will be deleted.

Notification:

ON:

This menu option allows you to activate/deactivate the "Notification" function. A green check mark and a beep informing when the target value (hysteresis) has been reached.

Hysteresis:

This menu option allows you to set the hysteresis from 0 to ± 2.0 K.

Sound:

Infinite: Beep until you press the "Back" key.

Beep: Single beep.



PUMP

Limits:

In "Limits" menu option, you are allowed to set the maximum and minimum speed to the pump, and also the maximum pressure.



TIMER:

Set:

You can set a target time (duration) in this menu option. When device functions are started normally, this time is displayed on the working screen. The device functions stop automatically once this time has expired. The running time is then displayed again on the display.

Note: To deactivate the target time specification, set the target time to 00:00.

Display:

Activate the timer display on the main display (working screen).



MODE:

Operating Mode A:

After power-on/power failure no automatic restart of functions.

Operating Mode B:

After power-on/power failure automatic restart of functions, depending on previous settings.

Operating mode C:

Set values (set in A or B) cannot be changed.

After power-on/power failure automatic restart of functions, depending on previous settings.

Operating mode D:

Confirmation request for set value changes, if functions are active. After power-on/power failure no automatic restart of functions.



DISPLAY:

Main Screen:

In the "Main Screen" menu option, you can specify what information will be displayed on the screen.

Device Number:

In the "Device number" menu option, you are allowed to edit the device number from 0 to 99 that will be shown on the main screen.

When using multiple units, it may be helpful to identify WiCo and station.

Note: Labeling device number to the station.



GRAPH

In this menu, you can set the options for the time-temperature diagram.

Automatic:

The scaling of the temperature axis (Y-axis) is automatically determined, depending on the target temperature and the actual internal and external temperature.

The time axis (X-axis) is permanently scaled to 30 minutes.

Manual:

Axis Assignment:

The temperature values to be displayed can be selected.

Axis scaling:

Scaling of the time (X) and temperature axis (Y) can be selected or set.



PROGRAMS:

Under programs, 10 user-defined temperature-time profiles can be created. A program can consist of up to 10 segments.

Once a program has been selected, the following options are available:

Start:

Starts the program upon request of loop mode.

Infinite loop: Upon completion of the last segment, the program continues with the first segment until you end the program by stopping a device function.

Loop Count: Indicates the total number of loop cycles until program end.

Note: At the end of the program all device functions are switched off.

Edit:

Edit/change program parameters.

Seg No.: Segment number.

Ctrl. Sensor (int/ext): Determines whether control is through the internal (int) or external (ext) temperature sensor.

Temp.: Target temperature.

Ctrl.Mode (Time / +/- x.xx K): In "Ctrl.Mode Time" the target values and settings of the segment are valid for the duration indicated in the column "Time hh:mm".

Afterwards, the next program segment is automatically executed.

In "Ctrl.Mode +/- x.xx K", the hysteresis (tolerance) of the actual temperature to the target temperature is set (e.g. +/- 0.1 K). The target values and settings of the segment are valid until the actual temperature reaches the target temperature +/- hysteresis for the first time.

Afterwards, the next program segment is automatically executed.

Pump rpm: Target speed of pump.

M1 (ON/OFF): MODULES M1-Valve:

OFF: M1 valve in initial state

ON: M1 valve in inverted initial state.

Note: The initial state of the M1 valve is defined in "MODULES M1-Valve" in "Default Status" as "Open" or "Closed".

M2 (ON/OFF): MODULES M2-Output switch
OFF: M2-output switch in initial state.
ON: M2-output switch in inverted initial state.

Note: The initial state of the M2-Output switch is defined in "MODULES M2-Output" in "Default Status" as "Open" or "Closed".

Cool: Activate/deactivate the cooling function.

OFF: cooling function is deactivated.

ON: cooling function is activated.

Delete: Deletes the selection highlighted in yellow (segment or program).

Insert: Inserts a new program segment after the selected segment.

Save: Saves changes.

View:

Temperature-time overview displays for the program with segments of the selected program.

Note: If hysteresis is set as "Ctrl.Mode +/-x.xx K" for one or more segments in the program, the duration of the program cannot be determined.
Press and turn the knob (R) to display the segment details.

Cancel: Cancels the process.

Note: Once the program has been started, the program no., segment no. (active/total) and the remaining duration of the segment or hysteresis are displayed in the graph.



MODULES:

In "Modules" the multifunction connection (14, see Fig. 31) outputs and inputs can be configured.

The outputs M1 and M2 can be controlled via "PROGRAMS".

M1-Valve:

ON:

Switches the external valve to the active state (inverted initial state).

Note: In a started „Program“ the M1 segment settings have higher priority.

Default Status:

Define the default state (OFF) of the external valve as "Open" or "Closed". It depends on the valve type (normally open or closed).

Refill:

Use the external M1 valve to automatically check the level when operating with water.

Observe "Default Status" setting.

Start: Set the valve switch-on point (ON).

Stop: Set the valve switch-off point (OFF).

Note: Prior to using the "Refill" menu option, check the function of the buoyage.

The "Refill" menu option has higher priority than the program M1 segment settings.

Temperature (int):

Temperature-dependent control of the M1-valve.

By entering the boundaries of "Higher" and "Low", the temperature range is set in which the M1-valve can be opened.

If you activate a rule "Int / ext> Set Temperature" is an automatic control of the M1-valve dependence on the target temperature.

By the "Hysteresis" value (set temperature hysteresis) the control stability can be optimized.

Note: int / ext stands according to the selected control mode (internal (int) or external (ext)) for the actual temperature to be controlled.

M2-Output:

Alarm:

Activate the alarm output (switch contact).

Error: In case of an error the "M2-Output — Switch" is activated (ON, inverted initial state).

Warning: In case of a warning the "M2-Output — Switch" is activated (ON, inverted initial state)

When "Warning" is activated, "Error" is simultaneously activated (see section "Error code").

Default Status: Defines the default status (OFF) of the "M2-Output — Alarm" as "Open" or "Closed". It depends on the alarm type (normally open or closed).

Switch:

Activate the switch output.

ON: Switches output into the active state (inverted initial state).

Note: When the "Program" is started, the M1 segment settings have higher priority.

Default Status: Defines the initial state (OFF) of the switch output as "Open" or "Closed". It depends on the switch type (normally open or closed).

Temperature (int)

Temperature-dependent control of the M2-switch.

By entering the boundaries of "Higher" and "Low", the temperature range is set in which the M2-switch can be opened.

If you activate a rule "Int / ext> Set Temperature" is an automatic control of the M2-switch dependence on the target temperature.

By the "Hysteresis" value (set temperature hysteresis) the control stability can be optimized.

Note: int / ext stands according to the selected control mode (internal (int) or external (ext)) for the actual temperature to be controlled.

M3-Input:

Ext. Standby:

External standby input to stop the device functions Temper and Pump.

ON: Activate the ext. standby function. Device functions are stopped in the event of an inverted initial state (ON) at the input.

Default Status: Define the default status (OFF) of the input as "Open" (high level) or "Closed" (low level).



OUTGAS:

In the menu option, you are allowed to set the outgas speed from 1000 rpm to 4600 rpm and outgas interval from 10 seconds to 240 seconds.

This function can be used when filling external devices such as laboratory reactors.



SAFETY:

Time out:

In the menu option "Time out", you can determine a time limit in the event of a communication breakdown between the station and the WiCo. The station continues to work with the preset target values until the preset time value has expired. Following that, the station runs using the preset safety temperature and safety speed.

Note: The initial time out is 30 seconds and you can define up to 60 minutes for this time limit.

Speed:

In the "Speed" menu option, you can specify a appropriate and safe speed for specified temperature.

Note: The factory setting of the safe speed is 1000 rpm and is activated after the time limit is set (see "Time out").

Temperature:

In the "Temperature" menu option, you can specify a temperature that is appropriate and safe for the circulating task.

Note: The initial safe temperature is 30 °C and is activated after the time limit is set (see "Time out").

Password:

In the "Password" menu, the menu settings can be locked by a 3-digit password.

Safe Temp. Confirmation:

This menu can confirm (safety temperature of the station) at start-up of the set "Safe Temp". A check mark (✓) indicates that the function is activated.



SETTING:

Languages:

The "Language" menu option allows you to select the desired language.

Display:

The "Display" menu option allows you to change the background color and brightness of the working screen. You can also activate the "Firmware Update Info" screen in this menu.

Sound:

The "Sound" menu option allows you to activate/deactivate the key tone.

Factory Settings:

Select the "Factory settings" menu option by turning and pressing the rotary/push knob. The system requests confirmation to restore the factory settings. Pressing the "OK" button resets all the system settings to the original standard values set at dispatch from the factory (see "Menu structure").

Communication:

The "Device name" menu option allows you to edit the device name. This can relatively identify the paired station. The "Bluetooth®" menu option allows you to activate/deactivate the "Bluetooth®" function. A check mark shows that the option is activated.

Information:

The "Information" menu option offers you an overview of the most important system settings of the device.

Interface and output

The device can be connected to a PC and operated with the laboratory software labworldsoft® through the RS 232 interface (11), USB interface (12) or USB interface on the WiCo.

Note: Please comply with the system requirements together with the operating instructions and help section included with the software.

USB interface:

The Universal Serial Bus (USB) is a serial bus for connecting the device to the PC. Equipped with USB devices can be connected to a PC during operation (hot plugging). Connected devices and their properties are automatically recognized. Use the USB interface in conjunction with labworldsoft® for operation in "Remote" mode and also to update the firmware.

USB device drivers:

First, download the latest driver for **IKA®** devices with USB interface from:

<http://www.ika.com/ika/lws/download/usb-driver.zip>.

Install the driver by running the setup file. Then connect the **IKA®** device through the USB data cable to the PC.

The data communication is via a virtual COM port. Configuration, command syntax and commands of the virtual COM ports are as described in RS 232 interface.

RS 232 interface:

Configuration

- The functions of the interface connections between the stirrer machine and the automation system are chosen from the signals specified in EIA standard RS 232 in accordance with DIN 66 020 Part 1.

- For the electrical characteristics of the interface and the allocation of signal status, standard RS 232 applies in accordance with DIN 66 259 Part 1.
- Transmission procedure: asynchronous character transmission in start-stop mode.
- Type of transmission: full duplex.
- Character format: character representation in accordance with data format in DIN 66 022 for start-stop mode. 1 start bit; 7 character bits; 1 parity bit (even); 1 stop bit.
- Transmission speed: 9600 bit/s.
- Data flow control: none
- Access procedure: data transfer from the stirrer machine to the computer takes place only at the computer's request.

Command syntax and format:

The following applies to the command set:

- Commands are generally sent from the computer (Master) to the device (Slave).
- The device sends only at the computer's request. Even fault indications cannot be sent spontaneously from the device to the computer (automation system).
- Commands are transmitted in capital letters.
- Commands and parameters including successive parameters are separated by at least one space (Code: hex 0x20).
- Each individual command (incl. parameters and data) and each response are terminated with CR LF (Code: hex 0x0d hex 0x0a) and have a maximum length of 80 characters.
- The decimal separator in a number is a dot (Code: hex 0x2E).

The above details correspond as far as possible to the recommendations of the NAMUR working party (NAMUR recommendations for the design of electrical plug connections for analogue and digital signal transmission on individual items of laboratory control equipment, rev. 1.1).

The NAMUR commands and the additional specific **IKA®** commands serve only as low level commands for communication between the device and the PC. With a suitable terminal or communications programme these commands can be transmitted directly to the circulator equipment. The **IKA®** software package, Labworldsoft®, provides a convenient tool for controlling circulating equipment and collecting data under MS Windows, and includes graphical entry features, for pump motor speed ramps for example.

Commands

Commands	Function
IN_PV_1	Read the external actual temperature
IN_PV_2	Read the internal actual temperature
IN_PV_3	Read the safety actual temperature
IN_PV_4	Read the pump actual speed
IN_SP_1	Read the internal setting temperature (if 0: internal control) Read the external setting temperature (if 1: external control)
IN_SP_3	Read the safety setting temperature
IN_SP_4	Read the pump setting speed

IN_TMODE	Read temperature control: 0: internal control 1: external control
OUT_SP_1 xxx	Set the internal setting temperature XXX (if 0: internal control) Set the external setting temperature XXX (if 1: external control)
OUT_SP_12@n	Set the WD safety temperature with echo of the set (defined) value.
OUT_SP_4 xxx	Set the pump speed XXX
OUT_SP_42@n	Set the WD-safety speed with echo of the set (defined) value.
OUT_TMODE_0	Set to Internal temperature control
OUT_TMODE_1	Set to External temperature control
OUT_WD1@n	Start the watchdog mode 1 and set the time for the watchdog to n (20...1500) seconds. Echo of the Watchdog time. During a WD1-event, the tempering and pump functions are switched off. This command needs to be send within the watchdog time.
OUT_WD2@n	Start the watchdog mode 2 and set the watchdog time to n (20...1500) seconds. Echo of the watchdog time. During a WD2-event, the set temperature is changed to the WD safety temperature and the pump set speed is set to the WD safety speed. This command needs to be send within the watchdog time.
RESET	Reset the PC control and stop the device functions.
START_1	Start the tempering function and stop the device functions.
START_4	Start the pump function
STOP_1	Stop the tempering function
STOP_4	Stop the pump function

Connections between device and external devices:

PC 1.1 Cable:

This cable is required to connect RS 232 interface (11) to a PC.




Fig. 28

Connection WiCo to station:





Fig. 29



Connection the device to PC:










Fig. 30

Multifunction interface:



- 1 M1 Output Valve + (+24Vdc/max. 0.8A)
- 2 M1 Output Valve -
- 3 M2 Output Alarm/Switch 1 (max. 30Vdc/ac/max. 1A)
- 4 M2 Output Alarm/Switch 2
- 5 M3 Input standby + (+5V ca. 10mA)
- 6 M3 Input standby - (0V only for Standby)
- 7 --- (reserved for later use, do not connect!)
- 8 --- (reserved for later use, do not connect!)

Fig. 31

Maintenance and cleaning

To avoid contamination, it is essential to check and change the bath fluid from regularly.


If water is used for bath fluid, we recommend to add Water bath protective media. The protective media stops the growth of algae, bacteria and other microorganisms with its bactericidal effectiveness. It protects the bath fluid and provides a long period of use for the fluid.

To keep the full cooling performance, the dust filter of the cooler must be checked regularly and cleaned if necessary.

- Switch off the device and disconnect mains cable.
- Open the front venting grid (5).
- Remove the bracket.
- Clean condenser dust filter with a vacuum cleaner or wash the filter in the water and dry it before assembly.

Note: Don't touch the condenser surface with hard parts.

Cleaning:



Use only cleaning agents which have been approved by **IKA®** to clean the device.

Dirt	Cleaning agent
Dye	Isopropyl alcohol
Construction material	Water containing tenside / isopropyl alcohol
Cosmetics	Water containing tenside / isopropyl alcohol
Foodstuffs	Water containing tenside
Fuel	Water containing tenside

For materials which are not listed, please request information from **IKA®** application support.

Wear protective gloves while cleaning the devices.

Electrical devices may not be placed in the cleansing agent for the purpose of cleaning.

Do not allow moisture to get into the device when cleaning.

Before using another than the recommended method for cleaning or decontamination, the user must ascertain with **IKA®** that this method does not damage the device.

Spare parts order:

When ordering spare parts, please give:

- machine type
- serial number, see type plate
- item and designation of the spare part, see **www.ika.com**
- Software version.

Repair:

Please send in device for repair only after it has been cleaned and is free from any materials which may constitute a health hazard.

For repair, please request the "Decontamination Certificate" from **IKA®**, or download printout of it from the **IKA®** website **www.ika.com**.

If you require servicing, return the device in its original packaging. Storage packaging is not sufficient. Please also use suitable transport packaging.

Error codes

Any malfunctions during operation will be identified by an error message on the display.

Proceed as follows in such cases:

- ☞ Switch off device using the main switch at the back of the device
- ☞ Carry out corrective measures
- ☞ Restart device

Error code	Effect	Cause	Solution
Error 01	Pump off Heating/refrigerating off	No external sensor	- Check this sensor
Error 02	Pump off Heating/refrigerating off	Motor over current (rate current)	- Reduce pump motor speed - Use fluid with lower viscosity - Check if the pump impeller is blocked
Error 03	Pump off Heating/refrigerating off	Motor over current (Max current)	- Reduce pump motor speed - Use fluid with lower viscosity - Check if the pump impeller is blocked
Error 04	Pump off Heating/refrigerating off	Motor hall signal missing	- Reduce pump motor speed - Use fluid with lower viscosity - Check if the pump impeller is blocked
Error 05 ❶	Pump off Heating/refrigerating off	Too high liquid level (when the heating function is activated)	- Check the fluid level and buoyage
Error 06 ❶	Pump off Heating/refrigerating off	Too low liquid level (when the heating function is activated)	- Check the fluid level and buoyage
Error 07	Pump off Heating/refrigerating off	Too high voltage	- Check the mains power
Error 08	Pump off Heating/refrigerating off	Too low voltage	- Check the mains power
Error 09	Pump off Heating/refrigerating off	Device internal temperature is too high	- Check the ambient temperature and let the device cool down - Check the fan and clean the grids at the rear side
Error 10	Pump off Heating/refrigerating off	PC communication failure	- Check communication cable
Error 11	Pump off Heating/refrigerating off	Temperature difference between control sensor and safety sensor is too much	- Check safety temperature circuit and bath fluid
Error 12	Pump off Heating/refrigerating off	Safety temperature alarm	- Check the bath temperature measurement
Error 13	Pump off Heating/refrigerating off	Heater switched off by safety circuit	- Check safety temperature set value, fluid level
Error 14	Pump off Heating/refrigerating off	Fan error	- Restart the device
Error 15	Pump off Heating/refrigerating off	Safety shutdown of the compressor	- Restart the device
Error 16	Pump off Heating/refrigerating off	Errors at the compressor (suction side)	- Restart the device
Error 17	Pump off Heating/refrigerating off	Errors at the compressor (pressure side)	- Restart the device
Error 20	Pump off Heating/refrigerating off	Fan cooling system error	- Restart the device
Error 21	Pump off Heating/refrigerating off	Internal error	- Restart the device

Warning message	Effect	Cause	Solution
High fluid level warning ⓘ	Warning message and high level icon appear	Too high fluid level	- Check the fluid level and buoyage - Drain excess fluid from bath.
Low fluid level warning ⓘ	Warning message and low level icon appear	Too low fluid level	- Check the fluid level and buoyage - Add more fluid to bath.

ⓘ Note: When the heating function is not activated, the screen show only these warning information if fluid level is too high or too low.

When the heating function is activated, if the fluid level goes slowly up/down to the warning level, the screen show high/low fluid level warning message and icon.

If the fluid level is corrected, the warning message and high/low level icon disappear.

If the fluid level increases/decreases further, the screen show Error 5/Error 6.

If the actions described fails to resolve the fault or another error code is displayed then take one of the following steps:

- Contact the service department
- Send the device for repair, including a short description of the fault.

Accessories

Tubing and hoses:

LT 5.20	Metal hose (isolated M16 x 1)
LT 5.21	PTFE hose (isolated M16 x 1)
H.PUR.8	PUR tube (nominal width 8 mm)
H.PUR.12	PUR tube (nominal width 12 mm)
H.FKM.8	FKM tube (nominal width 8 mm)
H.FKM.12	FKM tube (nominal width 12 mm)

Tubing insulations:

ISO. 8	Insulation (8 mm)
ISO.12	Insulation (12 mm)

Valve:

MV 1	Magnetic valve
CO V 1	Closed pressure valve

Additional accessories:

PC 1.1	Cable (RS 232)
Pt 100.3	Temperature sensor
Labworldsoft®	

See more accessories on www.ika.com.

Technical data

station		
Operating voltage	VAC	230 ± 10 % / 100–115 ± 10 %
Frequency	Hz	50 / 60
Max. input power	W	3000 (230 VAC) / 1750 (115 VAC)
Working temperature range	°C	- 25 ... + 200
Operating temperature range	°C	- 25 ... + 200
Temperature stability – internal temperature control 70 °C, water (according to DIN 12876)	K	± 0.01
Temperature control		PID (Automatic / user setting)
Temperature measurement, absolute accuracy		
Internal (int) (adjustable by calibration)	K	± 0.2
External (ext) (adjustable by calibration)	K	± 0.2
External Pt 100.3 temperature sensor tolerance to DIN EN 60751 class A, ≤ ± (0.15 + 0.002 x T), e.g. at max. 100°C (adjustable by calibration)	K	± 0.35 (at 100°C)
Temperature setting		Knob on WiCo
Temperature setting resolution	K	0.1
Temperature display		TFT LCD on WiCo
Temperature display resolution	K	0.01
Classification according to DIN 12876-1		Class III (FL) suitable for flammable and non-flammable fluids
Safety circuit (adjustable)	°C	0 ... + 260
Safety temperature display		TFT LCD on WiCo
Heating capacity	W	2500 (230 VAC) / 1250 (115 VAC) / 945 (100 VAC)
Cooling capacity according to DIN 12876 (at 4600 rpm): + 20 °C + 10 °C 0 °C - 10 °C - 20 °C	W	350 (400 W at 3200 rpm) 320 (370 W at 3200 rpm) 270 (320 W at 3200 rpm) 190 (240 W at 3200 rpm) 80 (130 W at 3200 rpm)
Refrigerant		R134a 
Refrigerant quantity	g	230
Max. refrigerating system pressure	bar	20
Pump speed (adjustable)	rpm	1000 ... 4600
Max. pump pressure/suction	bar	0.61 / 0.45
Max. flow rate (at 0 bar)	l/min	31
Bath volume	l	5–7
Fluid maximum viscosity	mm²/s	50
Sub-level protection		Yes
Interface		USB, RS 232, multi-function interface
Permitted on-time	%	100
Protection class according to EN 60 529		IP 21
Protection class		I
Excess voltage category		II
Contamination level		2
Permitted ambient temperature	°C	+ 5 ... + 32
Permitted ambient humidity	%	80
Dimension (W x D x H)	mm	275 x 490 x 690
Weight	kg	40
Operation at a terrestrial altitude	m	max. 2000

WiCo		
Permitted on time	%	100
Max. communication distance (dependent on the building)	m	15
Dimensions (W x D x H)	mm	160 x 40 x 105
Weight	kg	0.3
Ambient temperature	°C	+ 5 ... + 40
Ambient humidity (relative)	%	80
Protection class according to EN 60529		IP 40
Interface		USB
RB 1 Battery pack		
Voltage	V	5
Charging capacity	mAh	2000
Charging time	h	4.5
Working time	h	15
Battery type		Lithium-polymer

❗ **Note:** Refrigerant must be disposed of in accordance with local and national regulations.

Note: Complies to EN61000-3-11 subject to conditional connection: $Z_{max} = 0.13 \Omega$. If necessary, consult your electricity supplier.

Subject to technical changes!

Warranty

In accordance with **IKA®** warranty conditions, the warranty period is 24 months. For claims under the warranty please contact your local dealer. You may also send the machine direct to our factory, enclosing the delivery invoice and giving reasons for the claim. You will be liable for freight costs.

The warranty does not cover worn out parts, nor does it apply to faults resulting from improper use, insufficient care or maintenance not carried out in accordance with the instructions in this operating manual.

Pump performance curve

Pump performance curve measured with water:

(Measurements done according DIN 12876-2 with water at 20°C; pump in a closed-loop circuit).




Fig. 32

IKA®-Werke GmbH & Co.KG

Janke & Kunkel-Str. 10

D-79219 Staufen

Tel. +49 7633 831-0

Fax +49 7633 831-98

sales@ika.de

www.ika.com