Manual Supplement

Manual Title: 80 Series V Users Supplement Issue: 1

Print Date: May 2004 Issue Date: 5/04
Revision/Date: Page Count: 4

evision/Date: Page Count: 4

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 80 Series V CD Rev. & Date: 5/2004 CD PN: 2101999



80 Series V Users Manual Supplement

Change #1

On page 9, Table 4, add the following entry to the section:

On page12, add the following table: Error Messages

. •	•			
bAtt	Replace the battery immediately.			
diSC	In the capacitance function, too much electrical charge is present on the capacitor being tested.			
EEPr Err	Invalid EEPROM data. Have Meter serviced.			
CAL Err	Invalid calibration data. Calibrate Meter.			
read	▲Test lead alert. Displayed when the rotary switch is moved to or from the mA or A position. Test lead alert. Displayed when the rotary switch is moved to or from the mA or A position. Test lead alert. Displayed when the rotary switch is moved to or from the mA or A position. Test lead alert. Displayed when the rotary switch is moved to or from the mA or A position. Test lead alert. Displayed when the rotary switch is moved to or from the mA or A position. Test lead alert. Displayed when the rotary switch is moved to or from the mA or A position. Test lead alert. Displayed when the mA or A position. Test lead alert. Displayed when the mA or A position. Test lead to or from the mA or A position. Test lead to			

5/04 1

Manual Supplement 80 Series V Users

On page 15, under Zero Input Behavior of True RMS Meter (87), replace the first paragraph and two bullets with following:

True Rms Meters accurately measure distorted waveforms, but when the input leads are shorted together in the AC functions, the meter displays a residual reading between 1 and 30 counts. When the test leads are open, the display readings may fluctuate due to interference. These offset readings are normal. They do not affect the Meterís AC measurement accuracy over the specified measurement ranges.

Unspecified input levels are:

- AC voltage: below 3 % of 600 mV AC, or 18 mV AC
- AC current: below 3 % of 60 mA AC, or 1.8 mA AC
- AC current: below 3 % of 600 μA AC, or 18 μA AC

On page 21, under Measuring Capacitance, add the following note:

Note

If too much electrical charge is present on the capacitor being tested, the display shows \diSCî.

On page 43, under *Specifications*, replace the Electromagnetic Compatibility with the following:

Electromagnetic Compatibility: In an RF field of 3 V/m total accuracy = specified accuracy + 50 counts.

2 5/04

80 Series V Users Manual Supplement

On page 44, Table 10, replace footnote 1, with the following:

1. Below 10% of range, add 12 counts.

On page 46, replace Table 12 with the following:

Table 12. DC Voltage, Resistance, and Conductance Function Specifications

Function	Range	Resolution	Accuracy	
			Model 83	Model 87
Ÿ	6.000 V 60.00 V 600.0 V 1000 V	0.001 V 0.01 V 0.1 V 1 V	± (0.1 % + 1) ± (0.1 % + 1) ± (0.1 % + 1) ± (0.1 % + 1)	± (0.05 % + 1) ± (0.05 % + 1) ± (0.05 % + 1) ± (0.05 % + 1)
 m∨	600.0 mV	0.1 mV	± (0.3 % + 1)	± (0.1 % + 1)
Ω	600.0 Ω 6.000 kΩ 60.00 kΩ 600.0 kΩ	0.1 Ω 0.001 kΩ 0.01 kΩ 0.1 kΩ	$\pm (0.4 \% + 2)^{1}$ $\pm (0.4 \% + 1)$ $\pm (0.4 \% + 1)$ $\pm (0.7 \% + 1)$	$\pm (0.2 \% + 2)^{1}$ $\pm (0.2 \% + 1)$ $\pm (0.2 \% + 1)$ $\pm (0.6 \% + 1)$
nS	$6.000 \text{ M}\Omega$ $50.00 \text{ M}\Omega^2$ 60.00 nS^1	0.001 MΩ 0.01 MΩ 0.01 nS	$\pm (0.7 \% + 1)$ $\pm (0.7 \% + 1)$ $\pm (1.0 \% + 3)^{2}$ $\pm (1.0 \% + 10)^{1}$	$\pm (0.6 \% + 1)$ $\pm (0.6 \% + 1)$ $\pm (1.0 \% + 3)^{2}$ $\pm (1.0 \% + 10)^{1}$

^{1.} When using the REL Δ function to compensate for offsets.

5/04 3

^{2.} Add 0.5 % of reading when measuring above 30 M Ω in the 50 M Ω range.

Manual Supplement 80 Series V Users

On page 48, Table 14, replace footnote 2 with the following:

2. AC conversions for Model 87 are ac coupled, true rms responding, and valid from 3 % to 100 % of range, except 400 mA range (5 % to 100 % of range) and 10 A range (15 % to 100 % of range).

On page 50, Table 17, under Duty Cycle Range, replace the Accuracy with the following:

Within \pm (0.2 % per kHz + 0.1 %) for risetimes < 1 μ s.

4 5/04