

TECHNICAL REPORT

Page 4 of 8



LABORATORY ADDRESS:

Level 4, Block N & O, Faculty of Medicine, Universiti Malaya
50603 Kuala Lumpur.

Tel: +603-79676670 Email: tidrec@um.edu.my

Website: www.tidrec.com



HEAD OF LABORATORY: Sazaly Abu Bakar, Ph.D., FASc

REPORT TS4-0425.1-R

Product Evaluation Study

Product Details

Product:	ProDetect® COVID-19 Antigen Rapid Test (Saliva)
Product code:	PR-CVDCAgS20
Lot number:	PR-CVDCAgS21JUL09
Manufacturer:	Medical Innovation Ventures Sdn. Bhd.
Requested by:	Medical Innovation Ventures Sdn. Bhd.
Address:	Level 4, Biopharmaceutical Block, IPHARM, NIBM, MOSTI, Block 5-A, Halaman Bukit Gambir, 11700 Gelugor, Penang, Malaysia
Contact number:	+60 14 919-1548
Email:	amali@mediven.com.co
Date of request:	10 th August 2021
Type of sample tested:	Saliva

Executive Summary

The evaluation study was performed to determine the performance of a self-test kit - the ProDetect® COVID-19 Antigen Rapid Test (Saliva) in detecting SARS-CoV-2 antigen from saliva samples. The testing was performed on 30 SARS-CoV-2 positive and 30 SARS-CoV-2 negative saliva samples (by real-time RT-PCR). The positive samples were samples with Ct values between 17.79 – 29.63. The present evaluation showed that the ProDetect® COVID-19 Antigen Rapid Test (Saliva) was able to detect SARS-CoV-2 antigen in all of the positive samples indicating 100% sensitivity. No false positives were detected among the negative samples, indicating the test 100% specificity in detecting SARS-CoV-2 antigen in saliva samples.

TECHNICAL REPORT



LABORATORY ADDRESS:

Level 4, Block N & O, Faculty of Medicine, Universiti Malaya
50603 Kuala Lumpur.

Tel: +603-79676670 Email: tidrec@um.edu.my

Website: www.tidrec.com



HEAD OF LABORATORY: Sazaly Abu Bakar, Ph.D., FASc

1. Purpose and scope

The present evaluation was performed to determine the performance of a self-test kit – the ProDetect[®] COVID-19 Antigen Rapid Test (Saliva) in detecting SARS-CoV-2 antigen from saliva samples. The evaluation is a partial requirement for the product recommendation of use by the Medical Device Authority (MDA) and the Ministry of Health Malaysia.

2. Materials and methods

2.1 Description of device and intended use

The ProDetect[®] COVID-19 Antigen Rapid Test (Saliva) is a rapid lateral flow chromatographic immunoassay for the qualitative detection of SARS-CoV-2 nucleocapsid antigen in human oral fluid specimens. It serves as an aid in the diagnosis of COVID-19 for individuals suspected of SARS-CoV-2 infection. It is intended for home use only.

In principle, when the specimen is added to the device and it contains SARS-CoV-2 antigen, the antigen will react with the SARS-CoV-2 nucleocapsid antibody-coated particles at the T line region of the test cassette. A coloured line will appear. If the specimen does not contain SARS-CoV-2 antigen, no coloured line will appear indicating a negative result. As a procedural control, a line will always appear at the C region to indicate that sample has been added and membrane wicking has occurred.

2.2 Panel of samples

The samples used were archived saliva samples obtained from COVID-19 patients. They have been in storage for about 1-2 months in -80 °C. These samples have been previously tested for SARS-CoV-2 using real-time RT-PCR with amplification Ct values of 17.79 – 29.63. A total of 30 SARS-CoV-2 positive and 30 SARS-CoV-2 negative samples were used for the present evaluation.

2.3 Detection of SARS-CoV-2 Antigen using the ProDetect[®] COVID-19 Antigen Rapid Test (Saliva)

Materials provided in the kit are the test device, the collection device, a biosafety bag, single-use buffer, a procedure card and an instructions for use. The test procedure was performed as follows:

- The test device was removed from its packaging pouch and placed on a level surface.
- About 50 µl of the saliva sample was mixed with 50 µl of the buffer. The mixture was mixed well and all of it was transferred to the test device sample well.
- The test device was incubated at room temperature for about 15-20 min following which the result was read. Results read after 30 min were considered invalid. The appearance of two

TECHNICAL REPORT

Page 6 of 8



LABORATORY ADDRESS:

Level 4, Block N & O, Faculty of Medicine, Universiti Malaya
50603 Kuala Lumpur.

Tel: +603-79676670 Email: tidrec@um.edu.my

Website: www.tidrec.com



HEAD OF LABORATORY: Sazaly Abu Bakar, Ph.D., FASc

lines at the T and C region indicated the presence of SARS-CoV-2 antigen in the tested sample and a negative result was indicated by the appearance of only one line at the C region.

All procedure was performed in a biosafety cabinet to minimize potential exposure to SARS-CoV-2.

3. Results

A total of 30 SARS-CoV-2 positive and 30 negative saliva samples were used for the evaluation. Performance of the product was calculated using the Evidence-based Medicine (EBM) Diagnostic Test Calculator (<https://ebm-tools.knowledgetranslation.net/calculator/diagnostic/>; Table 2). Results showed that the ProDetect[®] COVID-19 Antigen Rapid Test (Saliva) was able to detect SARS-CoV-2 antigen in all 30 positive samples tested resulting in a sensitivity of 100% (Table 1 & 2). No false positive was detected among the SARS-CoV-2 negative specimens indicating the product specificity of 100%.

Concordance analysis showed that there was a perfect agreement ($\kappa=1.000$; $p<0.001$) between the test product and the real-time RT-PCR used as the comparator assay.

Limitation of the evaluation

Due to resource and time constraint, the self-test product was not tested on field. Only archived saliva samples were used. Integrity of these specimens were ensured by their storage in -80 °C and minimizing the freeze-thaw cycles.

TECHNICAL REPORT



LABORATORY ADDRESS:

Level 4, Block N & O, Faculty of Medicine, Universiti Malaya
50603 Kuala Lumpur.

Tel: +603-79676670 Email: tidrec@um.edu.my

Website: www.tidrec.com



HEAD OF LABORATORY: Sazaly Abu Bakar, Ph.D., FASc

Table 1. Detection of SARS-CoV-2 antigen in saliva samples using the ProDetect[®] COVID-19 Antigen Rapid Test (Saliva).

		Real-time RT-PCR		Total
		Pos	Neg	
Tested Product	Pos	30	0	30
	Neg	0	30	30
Total		30	30	60

Table 2. Analysis of SARS-CoV-2 antigen detection using the ProDetect[®] COVID-19 Antigen Rapid Test (Saliva) on saliva samples.

Pos	Neg	Sensitivity (95% CI) ^a	Specificity (95% CI)	PPV ^b (95% CI)	NPV ^c (95%CI)	Concordance (Kappa value, κ)
30	30	100% (88.6-100)	100% (88.6-100)	100% (88.6-100)	100% (88.6-100)	1.000 (p<0.001)

^a Confidence Interval; ^b Positive predictive value; ^c Negative predictive value

TECHNICAL REPORT



LABORATORY ADDRESS:

Level 4, Block N & O, Faculty of Medicine, Universiti Malaya
50603 Kuala Lumpur.

Tel: +603-79676670 Email: tidrec@um.edu.my

Website: www.tidrec.com





HEAD OF LABORATORY: Sazaly Abu Bakar, Ph.D., FASc

4. Conclusion

Evaluation of the ProDetect® COVID-19 Antigen Rapid Test (Saliva) was successfully completed. The tested product showed that it was able to detect the presence of SARS-CoV-2 antigen in all of the tested saliva samples (100% sensitivity) with 100% specificity.

Disclaimer

This report (including any disclosures and attachments) has been prepared for the exclusive use and benefit of the Client/Requestor and solely for the purpose for which it is provided. We do not provide any warranty on the report nor do we accept any liability if findings in this report is not reproducible or devoid of any errors, or used for an alternative purpose from which it is intended, nor to any third party in respect of this report.

PREPARED BY:	
 Juraina Binti Abd. Jamil Research Officer Tropical Infectious Diseases Research & Education Centre (TIDREC) Higher Institution Centre of Excellence (HiCoE) Universiti Malaya 50603 Kuala Lumpur	 Teoh Boon Teong, PhD Senior Lecturer Tropical Infectious Diseases Research & Education Centre (TIDREC) University of Malaya 50603 Kuala Lumpur, Malaysia
Name: Juraina Abd Jamil, MMedSc	Name: Teoh Boon Teong, PhD
Position: Deputy Technical Manager	Position: Technical Manager
Date: 10 th August 2021	Date: 10 th August 2021

APPROVED BY:
 Sazaly Abu Bakar, PhD, FASc Professor & Director Tropical Infectious Diseases Research & Education Centre (TIDREC) Higher Institution Centre of Excellence (HiCoE) University of Malaya 50603 Kuala Lumpur, Malaysia
Name: Sazaly Abu Bakar, PhD, FASc
Position: Director
Date: 10 th August 2021