

# Manual Supplement

Manual Title: 1507/1503 Users  
Part Number: 2416013  
Print Date: June 2005  
Revision/Date:

Supplement Issue: **2**  
Issue Date: 8/05  
Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 4, under **Battery Saver (Sleep Mode)**, replace the last sentence in the first paragraph with the following:

The Tester comes out of Sleep mode when when the rotary switch is turned to off and back on.

On page 24, under Model 1503, replace the Display Range, Output Voltage 500V:

From: 0.1 to 20.00 M $\Omega$

To: 0.01 to 20.00 M $\Omega$

## Change #2

On page 5, under Table 2, add the following:

### Warning

To avoid shock, injury, or damage to the Tester, use only accessories with 1000V CAT III or 600V CAT IV rating.