



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

NIOSH Reference: TN-18158
Mfr. Reference: MAK-1161

Centers for Disease Control
and Prevention (CDC)
National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
P.O. Box 18070
Pittsburgh, PA 15236-0070
Phone: 412-386-4000
Fax: 412-386-4051
December 5, 2011

Mr. Alexander Freedman, President
Makrite Industries Inc.
236 Upland Avenue
Newton Highlands, Massachusetts 02461-2003

Dear Mr. Freedman:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted November 15, 2011. This request was for an extension of approval to TC-84A-5411 and TC-84A-5463 to add private label versions of these N95 filtering facepiece models 9500-N95 and 9500-N95S, for Quebee Safety Technologies (Int'l) PTE, LTD, of Singapore, as the models QB-910I-N95 and QB-910IS-N95, reference the assembly matrices included in MAK1161AM1.xls.

This request is granted. Approval is granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English.

The CD enclosed with this letter contains the final respirator labels. The abbreviated labels have been accepted as submitted. The cautions and limitations which apply to this approval are on the approval labels. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.


The approved assemblies consist of the parts as listed on the approval labels and the assembly matrices. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

The manufacturer is responsible for properly packaging, labeling, and controlling the respirators produced under this private label approval. At a minimum, the items that must be controlled are the approved user instructions, all approval labeling, all approved packaging, use claims, marketing materials, and the respirator design and construction details. Any change to these NIOSH-approved respirators or to the approval documentation without prior notification and approval is a violation of this approval and renders this certification as invalid.

Page 2 – Mr. Alexander Freedman – TN-18158

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely yours,

A handwritten signature in dark ink, appearing to read "Heinz W. Ahlers", is written over the typed name.

Heinz W. Ahlers
Chief, Technology Evaluation Branch
National Personal Protective Technology Laboratory

Enclosures



National Institute for Occupational Safety and Health
National Personal Protective Technology Laboratory
Technology Evaluation Branch
Certification, Evaluation and Testing Section
P.O. Box 18070
Pittsburgh, PA 15236

PROJECT BACKGROUND REPORT

Task Number: TN-18158

Manufacturer: Makrite Industries, Inc.

Prepared by: Tom Pouchot

Date: November 30, 2011

Background Information

In an application accepted November 15, 2011 Makrite Industries requested extensions of approvals to TC-84A-5411 and TC-84A-5463 to add private label versions of these N95 filtering facepiece, models 9500-N95 and 9500-N95S, for Quebee Safety Technologies (Int'l) PTE, LTD, of Singapore, as the models QB-910I-N95 and QB-910IS-N95, reference the assembly matrices included in MAK1161AM1.xls.

No testing was performed as the addition of the private label versions did not change the respirator designs or functions. These respirators were successfully tested under TN-17685 and TN-17811.