

Certificate of compliance

Date of certificate: 2021 - 02 - 15

Manufacturing date

2020/11/25

Page: 1/1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE

Product description

SelfstandingTube 50mL PPGraduated G reenCap Sterile25Bag500CS

Catalogue number

TU50-013

Lot number

754603

Lots: 754603202011

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-3

Expiration date

2023/10/01

Expiration statement

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

Cédric MAES

Plant Quality Manager, Borre

maesc@corning.com