



422 Aldo Ave., Santa Clara, CA 95054

Product Certificate

Thermo Scientific
Fiberlite Products

This information is the product certification for item #

010-1456

Description:

Thermo Scientific Fiberlite Superspeed 1L Bottle PPCO Assy (sets of 2)

Lot#:

20411251400485

Issued Date

4/11/2025

Manufacture Date

4/11/2025

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI
010-1361	Thermo Scientific Fiberlite PPCO Bottle 1000 mL	Component Part	-	-	-
	Resin used for Bottle	POLYPROPYLENE COPOLYMER	6345	Passed	Passed
010-1457	Thermo Scientific Fiberlite Bottle 1000 mL Closure	Component Part	-	-	-
	Metal used for closure	Anodized Aluminum Alloy	-	N/A	N/A
Part of 010-1361	Thermo Scientific Fiberlite Bottle 1000 mL Closure liner	Component Part	-	-	-
Part of 010-1361	Thermo Scientific Fiberlite Bottle 1000 mL O-ring	Component Part	-	-	-
	RESIN, PP,20%GF,NAT,INJ	PP, 20% GLASS REINFORCED	-	Passed	Passed
	COLOR, WHT, MULTI	COLORANT, WHITE	16513	Passed	Passed

Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity and USP Class VI. If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol. If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5. This sealing closure contains a silicone gasket. The gasket material meets requirements of CFR 21, Section 177.2600 of the Federal Food and Drug Act. The material has also been tested and shown to comply with USP Class VI requirements and has been shown to be non-cytotoxic.

ANIMAL DERIVED MATERIALS (BSE/TSE) - All resins and colorants used at the manufacturing site comply with the latest revision of EMA/410/01 section 6.4. The purchased component(s) used in this product also comply with the latest revision of EMA/410/01 section 6.4.