

SPECIFICATIONS: VEIN FINDER/VIEWER Qty: 1 unit

Description:

1. The product should be able to find veins for the purpose of Phlebotomy.
2. It should have the color adjustment option according to the different people's skin tone.
3. This vein finder should have different finding modes based the depth of vein including child mode.
4. It should be light weight.
5. It should be small and easy to be hand held.
6. The viewer should be stand mountable and hand held.
7. The stand should be supplied along with vein finder.
8. The veins should stand out in the light background.
9. It should have an inverse display, with veins appearing lighter compared to background.
10. The battery should last at least 3 hours.
11. It should be supplied with the battery charger or charging station.
12. The battery status should be displayed on the screen.
13. The Viewer should function at temperatures of 10° C to 50° C.

Standards and Requirements:

- It should comply to IEC/EN 60825-1 laser safety norms.
- USFDA, CE and or ISO certifications or relevant standards certification.
- European CE Certified under PPE (Personal Protection Equipment) directive 89/686/EEC.
The valid certificate from Notified body of EU with the Notified body no. on the certificate shall be enclosed.

Additional Remarks or Requirements:

- Complete product details to be enclosed with the original brochure or catalog. (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, consumables and minimum supplies to be stated clearly.

