

General Surgical Instruments

S.NO	INSTRUMENT NAME	Qty (nos)
1	Forster Sponge Holder 18cm	10
2	Bergmann Sponge Holder 24 cm	20
3	Halsted Mosquito haemostatic forceps curved 12 cm	50
4	Halsted Mosquito haemostatic forceps curved 14 cm	20
5	Allis tissue grasping forceps 15.5 cm	70
6	Babcock tissue grasping forceps 15.5 cm	10
7	Backhaus towel clamps 13cm	15
8	kidney dish 250x140x40 mm	5
9	Scalpel no 3	20
10	Scalpel no 4	20
11	Metzenbaum dissecting scissor curved 25 cm	15
12	Metzenbaum dissecting scissor curved 18 cm	15
13	Dissecting forceps 14.5 cm	10
14	Adson dissecting forceps 15 cm	15
15	Tissue forceps (single toothed)13cm	15
16	Tissue forceps (multiple toothed)14.5 cm	2
17	Lister sinus forceps 16 cm	5
18	Langenbeck Mini retractor 16 cm (tip 21x8mm)	6
19	Langenbeck retractor 22 cm (tip 50x11mm)	62
20	Deaver retractor 30.5cm (tip 25mm)	2
21	Morris retractor 25cm	2
22	Joll self retaining retractor 15.5 cm	10
23	Kelly dissecting and ligature forceps 21.5cm	6
24	S S Bowl 80x40mm	10
25	S S Instrument tray 350x240x40mm	3
26	S S Instrument tray 470x320x50mm	4
27	Wire Cutter 23 cms	2
28	Fistula Probe with eye and maleable	2
29	Lockhart -Mummery fistula probe curved 16.5 cm	2
30	Kelly Proctoscope 25x75mm	2

62	Dissecting forceps 14.5 cm	10
63	Adson dissecting forceps 15 cm	15
64	Adson dissecting forceps TC GOLD 15cm Non tooth	15
65	cushing dissecting forceps 17.5 cm	10
66	Satinsky vascular clamp15.5 cm	4
67	Satinsky vascular clamp20 cm	4
68	Mayo-Robson Intestinal Noncrushing Clamp 26 cms straight	4
69	Mayo-Robson Intestinal Noncrushing Clamp 25 cms Curved	4
70	Cooley Suction Tip 31 cms, 10 mm tip	3
71	Wire Twisting Forceps 19 cms TC Gold	4
72	Heiss Right Angle Hemostatic forceps 20cms	4
73	Varicose Vein Stripper	2
74	Humby dermatome 32 cm	2
75	Santulli Bull Dog clamp straight 7 cm	4
76	De bakey bull dog clamp straight 8.5 cm	4
77	De bakey bull dog clamp straight 12 cm	4
78	Goldmann - Fox Iris scissor TC Gold 13.5 cm	5
79	Mayo Hegar needle holder TC GOLD 20 cm	10
80	Bozemann/Wertheim needle holder Tc Tipped 20.5 cms	8
81	Micro scissor (Ti Tipped) 15 cm, curved	10
82	Micro scissor (Ti Tipped) 15 cm, straight	10
83	Adson toothed microforceps 12 cm	5
84	Mayo dissecting Scissor 14.5cm	10
85	Rings for hemarrhoidal Ligator	2 PKT
86	Heager cervical dialator(4,6,8,10,12,14,16,18 mm)	2
87	SS drum Large 15x 12 inches	10
88	SS drum Large 11x 9 inches	10
89	Mixer Baby Hemostatic Forceps 13cm	6
90	Sims Bivalve Speculum Small	6
91	Cusco Vagainal Speculum Medium	6
92	Cusco Vagainal Speculum Large	6



Technical Specification for Surgical Instruments

Technical Specifications mentioned below are of minimum parameter, products offered must meet these or exceed all requirements herein.

GENERAL SPECIFICATION

- All instruments should be of international quality and made from surgical grade stainless steel Documentary evidence required for grade of material.
- The all instruments should ensure in withstanding repeat sterilization without any deterioration to strength, edge quality or surface degradation.
- The instruments should be guaranteed against metal fatigue and rust for at least 02 years. Further repair should be available for next 5 years.
- The brand name along with catalogue number should be etched on the instruments.
- Submission of a Satisfactory Performance Certificate from a Government Hospital is considerable.
- Prior demonstration of instruments for approval if needed. It is the sole discretion of the department to approve or disapprove the quality.
- European CE & US FDA approved. Validation certificates must be submitted during technical bid.
- The medical instruments have a quality management system in accordance with the standard ISO 13485.
- Material Grade- AISI 410 and 420 certified.

ADDITIONAL CONDITIONS

- Bidder must enclose original literatures & technical data sheet in the support of the technical bid.
- Physical demo should be arranged at the time of requirement.
- The bidder should ensure the guarantee that any instrument will be free of defects in materials and workmanship, when used under normal conditions for its intended purpose. In the unlikely event that any instrument is found to be unsatisfactory in normal use within 2 years of purchase it will be refurbished or replaced free of charge.
- Technically Qualified Bidders may be asked to submit a few samples their quoted items for physical verification.
- GRH shall have the right to inspect and/or to test the goods to confirm their conformity to the Specifications
- After the goods arrival at the final destination, GRH shall be the final authority to reject full (or) any part of the supply which is not confirming to the specification and other terms and conditions.



- No payment shall be made for rejected Stores. Rejected items must be removed by the Bidders within two weeks of the date of rejection at their own cost and replaced immediately.
- Companies who have relevant experience in successful execution of similar work in at least 2 government will be preferred.
- Company should provide material certificates.
- Instruments should carry Ledger marking indicating the Manufacturer's name and Batch No. or Lot No.

