



ހުލުހުމާލެ ޖެނެރަލް ހޮސްޕިޓަލް ގެ ފަންޖަރުކުރުމުގެ ސަރަޙައްދުގައި
ޖެނެރަލް ހޮސްޕިޓަލް ގެ ފަންޖަރުކުރުމުގެ ސަރަޙައްދުގައި
ޖެނެރަލް ހޮސްޕިޓަލް ގެ ފަންޖަރުކުރުމުގެ ސަރަޙައްދުގައި

TENDER/HMH/2024/008

FEBRUARY 11, 2024
HULHUMALE 'HOSPITAL
Huvandhuma Hingun, Hulhumale', Rep. of Maldives


ANNEX1: SPECIFICATION SHEET

SECTION 1

GENERAL INFORMATION

NAME	Pediatric Treadmill
CATEGORY	Medical Equipment
CLINICAL PURPOSE	It is used for patients after stroke, patients in long-term bed and other lower limb dysfunction. With the help of the gait training device, the patients could recover the lower limb function gradually, the muscle becomes stronger and reduce stiffness

TECHNICAL SPECIFICATION

TECHNICAL SPECIFICATION	<ol style="list-style-type: none"> 1. Material: Aluminium 2. Height: 180-222cm 3. Adjustable height range of armrest: 75-158cm 4. Adjustable front to back range of armrest: 16-30cm 5. Width of armrest: 59cm 6. Loading Weigh: 80kg 7. Dimension: 116 x 85 x (170-210) cm 8. Speed: 0-11 MPH 9. Gradient of treadmill: 0%-15% 10. Column Lifting adjustment: 0-45cm 11. Gait training power adjustment: 0-120N 12. Power: <1.8kw 220v 50Hz DC24V 
--------------------------------	--

SECTION 3

GENERAL INFORMATION

NAME	Hippotherapy Machine
CATEGORY	Medical Equipment
CLINICAL PURPOSE	Assessment tool required in physiotherapy and occupational therapy that effectively strengthens the spine and helps to maintain a good posture, develop muscles strengthen and flexibility, trains the nervous system to improve motor functions and normalize muscle tone.

TECHNICAL SPECIFICATION

TECHNICAL SPECIFICATION	<ol style="list-style-type: none"> 1. Minimum user weight: 100 KG 2. Must have rubber feet towards the floor 3. Must include a visible display to show the selected type of workout and movement. 4. Must include wheels and handles to move the saddle around short distances 
--------------------------------	---