



Procurement Notice

1. Introduction

The Government of Maldives, through the Ministry of Health, is seeking to procure medical equipment for the Dr. Abdul Samad Memorial Hospital, GDH-Thinadhoo, Maldives. The procurement is identified by the reference number (IUL)ASMH-Procure/ASMH/2024/22.

The procurement is open to all qualified bidders. The estimated value of the procurement is approximately MVR 1,000,000.00 (One Million Maldivian Rufiyaa).

The procurement is divided into two lots:

- Lot 1: Medical equipment for the hospital.
- Lot 2: Medical equipment for the hospital.

The procurement is open until the closing date and time specified in the tender documents. The procurement is subject to the terms and conditions of the tender documents.

2. Tender Documents

The tender documents are available for download from the website tender@asmh.gov.mv. The tender documents include the following information:

- Reference number: 6845160
- Estimated value: MVR 1,000,000.00
- Closing date and time: 25th August 2024, 13:00

The tender documents are available in both English and Maldivian. The procurement is open to all qualified bidders.

2.2. Tender Documents

- Reference number: 6845160
- Estimated value: MVR 1,000,000.00
- Closing date and time: 25th August 2024, 13:00

2.3. Tender Documents

- Reference number: 6845160
- Estimated value: MVR 1,000,000.00
- Closing date and time: 18th August 2024, 14:00

3. Contact Information

For more information, please contact the procurement officer at procurement@asmh.gov.mv.

Reference number: 6845160



18.3.2. 18.3.2. 18.3.2. 18.3.2. 18.3.2.

18.3.3. 18.3.3. 18.3.3. 18.3.3. 18.3.3.

18.3.4. 18.3.4. 18.3.4. 18.3.4. 18.3.4.

19. 19. 19. 19. 19.

19.1. 19.1. 19.1. 19.1. 19.1.

20. 20. 20. 20. 20.

20.1. 20.1. 20.1. 20.1. 20.1.

20.2. 20.2. 20.2. 20.2. 20.2.

20.3. 20.3. 20.3. 20.3. 20.3.

20.4. 20.4. 20.4. 20.4. 20.4.

20.5. 20.5. 20.5. 20.5. 20.5.



ސަލާމް ޖެނެރަލް ހޮސްޕިޓަލް

Equipment: Table Top Centrifuge - 28 Wells

1. Adjustable RPM from 50-15000
2. Adjustable RPM from 50-24400
3. Should be supplied with safety lid and lock.
4. Should have digital speedometer and timer.
5. Should have imbalance detector and automatic cutoff.
6. Should be CE Certified.
7. Should work on 200-240Vac 50Hz power supply
8. Should have capacity of 4*290ml (28 wells)
9. Should provide Maldivian standard power cord
10. Should be brushless type & maintenance-free motor with frequency drive
11. Should have programmable time and speed.
12. Should have individual Acceleration and deceleration stages
13. Should be low noise operation type
14. Should be Supplied with all Standard Accessories and Consumables for blood tube package.
15. Should provide service in Maldives
16. Spare parts, accessories, consumables should be available for minimum 10 years from the date of installation the machine
17. Should provide warranty card, calibration report, service, operation manuals and CDs
18. User training should be provided(onsite)

Safety Profile

Should be provided with terminal for good earth connection to preclude electric disturbances while recording-

- Must have a safety certificate or valid detailed electrical and functional safety test report from a recognized competent authority
- Copy of the certificate / test report shall be produced along with the technical bid.

Installation & Training

The firm should install the instrument at the designated location and provide onsite one-day training/demonstration of operation centrifuge.

Warranty

- Warranty should start from the date of installation
- Performance warranty of at-least 5 years from date of installation
- In case of breakdown of the machine, the supplier shall make the machine functional by repair (including replacement of parts) free of cost at the user site, within 3 (three) days of the receipt of complaint, or replace the machine (if necessary).

After Sales Services

- The suppliers should have adequate after sales service facilities covering all districts of the country.



ޔ. ސ. ޖެނެރަލް ސަކަރު ޑޮކްޓަރު ޅަބްދުލް ސަމަދު ފަސިޔަލް

Dr .Abdul Samad Memorial Hospital

GDH·Thinadhoo, Maldives ޔ. ސ. ޖެނެރަލް ސަކަރު ޖެނެރަލް ސަކަރު ޅަބްދުލް ސަމަދު ފަސިޔަލް

- They should have infrastructure and trained manpower to attend to any complaints within 3 days of receipt of complaints
- Should attend the preventive maintenance of machine during warranty as well as in contract once in every year