

ANNEX1: SPECIFICATION SHEET

GENERAL INFORMATION

NAME	WAVEFRONT ABERROMETER WITH HEIGHT ADJUSTABLE TABLE (Essilor WAM800 or equivalent)
CATEGORY	Medical / Ophthalmology Equipment
CLINICAL PURPOSE	Provides visual acuity simulation allowing a better understanding of the patient vision. It is a multifunctional diagnostic device combining wavefront aberrometer, corneal topography, retroillumination, tonometry, pupillometry, keratometry, pachymetry, eye fatigue, objective refraction (only 90 seconds for both eyes).

TECHNICAL SPECIFICATION

TECHNICAL SPECIFICATION	<p>1. Functional Requirements:</p> <p>Glaucoma Screening:</p> <ul style="list-style-type: none"> • Integrated tonometry for IOP measurement • Visual field analysis compatibility <p>Cataract Screening:</p> <ul style="list-style-type: none"> • High-resolution retroillumination for lens opacity detection <p>Objective Vision Evaluation:</p> <ul style="list-style-type: none"> • Automated refraction assessment with high accuracy <p>Patient Vision Simulation:</p> <ul style="list-style-type: none"> • Software for simulating various visual conditions for patient understanding <p>Monocular Accommodation Assessment:</p> <ul style="list-style-type: none"> • Dynamic accommodation range measurement <p>Efficiency in Optical Environments:</p> <ul style="list-style-type: none"> • Compact design for easy integration into clinical settings <p>2. Operational Requirements:</p> <ul style="list-style-type: none"> • Fully automated with touch-screen interface • Automatic alignment and measurement • Rapid Processing: <ul style="list-style-type: none"> • Complete diagnostic analysis within 2 minutes • User Interface: <ul style="list-style-type: none"> • Intuitive user interface with customizable settings • Integration with EHR systems • USB, Ethernet, and wireless connectivity options <p>3. Training:</p> <ul style="list-style-type: none"> • Comprehensive application and biomedical service training (overseas if required) <p>4. Warranty:</p> <ul style="list-style-type: none"> • Warranty and technical support must be provided for a minimum of 5 years.
--------------------------------	---

