

BIO-MEDICAL ENGINEERING SECTION
Addu Equatorial Hospital
TECHNICAL SPECIFICATION
T- Piece Resuscitation Set for GE Panda Warmer

Parameter	Minimum Requirement
Item	Disposable T-piece neonatal patient circuit with neonatal face mask
Reference	GE NF-157-0GE or compatible equivalent
Compatibility	Shall be compatible with the existing GE Panda Infant Warmer
Type	Disposable neonatal T-piece patient circuit
Mask	Neonatal face mask, Size 0 , suitable for newborn resuscitation
PEEP Control	Adjustable PEEP control valve
Oxygen/Flow	Suitable for controlled delivery of 21%–100% oxygen and free-flow oxygen
Tubing	Flexible, lightweight and crush-resistant tubing
Complete Set	Circuit and mask shall be supplied as a complete ready-to-use assembly
Compatibility Evidence	Bidder shall provide manufacturer's documentation confirming compatibility with the GE Panda warmer

Photo Reference:

