



## **ANNEX 1**

### **TECHNICAL SPECIFICATIONS: GASTRO VIDEOSCOPE**

#### **1. GASTRO VIDEOSCOPE (EVIS EXERA III (GIF-XP190N))**

- a. Channel Width: 2.2 mm
- b. Working Length: 110 cm
- c. Field of View: 140°
- d. Direction of View: Forward viewing
- e. Depth of Field: 3-100 mm
- f. Outer Diameter: 5.4 mm
- g. Outer Diameter Insertion Tube: 5.8 mm
- h. Max Angulation Up: 210°
- i. Max Angulation Down: 90°
- j. Max Angulation Right: 100°
- k. Max Angulation Left: 100°

#### **2. ACCESSORIES (TO BE QUOTED SEPERATELY)**

- i. Should quote accessories and any other items that may be required to effectively use the laser unit in clinical setting.

