





ANNEX-1

MEDICAL CONSUMABLE AND LABWARE REQUIREMENT & SPECIFICATIONS

Item no:	Item	Description	Required quantity
1	Microscope objectives for Philip Harris Microscopes (x 40, high power) MOTIC 40/ 0.65 NA :160/0.17	To be used with Philip harris microscope	10 no's
2	Microscope objectives for Philip Harris Microscopes (x 100, oil immersion) PLAN 100/1.25, NA: 160/0.17	To be used with Philip harris microscope	10 no's
3	Test tube racks (4 X 10 Format)	Racks snap together for easy assembly. Sturdy construction, stackable, with alphanumerical positions for clear sample identification. Operating range - 20°C to 90°C. Autoclavable at 121°C Hold max 20mm W test tubes	10 Racks



[Handwritten signature]

4	Universal Bottles, 50ml	50 ml, glass with aluminium screw cap and rubber seal. Spare caps available.	100 Bottles	
5	Universal Bottles, 25ml	25 ml, glass with aluminium screw cap and rubber seal. Spare caps available.	100 Bottles	

Handwritten signature in blue ink.