

ANNEX 1

Specifications

1. Haemodynamic Monitoring System:

Req. Quantity 1 unit

Description:

- **The monitor should be compatible with BRANSIST ALEXA (Type-F12) -SHIMADZU-Digital angiography system.**
- 18" or more color wave monitor (TFT, Preferably LED) for patient dialog and real time waveforms with programmable layout and digital monitoring readout... Two such monitor required. One TFT monitor in examination room with ceiling suspension and one in console.
- An 18" or more remote color wave form monitor, to be mounted in the examination room with video switch between patient dialog and real time waveforms
- The recorder system should be integrated with the angiographic system so that patient demographics once entered should automatically captured by the recording system
- One workstation for offline Angio viewing and recording.
- The following features of Hemodynamic recording should be available.
 - Maintain Full Patient File Record
 - 12 Lead ECG Amplifier with floating input
 - At least 2 pressures with floating inputs
 - Time and amplitude measurement with electronic calipers
 - Automatic store of all analyzed data.
 - Facility to measure oxygen saturation and Hb concentration
 - Report & Archive
 - System Setup
 - Built-in Database (SQL Server)
 - Laser Printer with minimum 16 MB memory with minimum 1200 dpi
 - Real Time Calculation of Fraction Flow Reverse.
 - Real – Time FFR Values upon onset of hyperemia
 - Storage of ECG/pressure recording on CD
 - User configuration setup.
 - Conversion of homodynamic reports into DICOM Compatible image data format



Required Accessories: -

- ECG cables and reusable pressure transducers
- Software should be provided for offline hemodynamic calculations such as cardiac output, gradients and stunt estimations.
- Hemodynamic monitoring system with all standard accessories for adult, pediatric and neonatal patients.

Standards and Requirements:

- USFDA approved product. All other accessories should be USFDA, CE and or ISO certifications or relevant standards certification.
- Electrical safety conforms to standards for electrical safety IEC-60601-General Requirements

Shall comply with AERB and BARC guidelines.

- Power requirement:250V, 50Hz.
- The supplier should visit the site and install the monitor. Also, the supplier should do the monitor interface with the Cathlab console system.
- User/Application training to end users and In-house engineers should be provided by authorized trainers after the installation.

Additional Remarks or Requirements:

- Complete product details to be enclosed with the original brochure or catalogue (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, and minimum supplies to be stated clearly.



2. Ceiling Mounted X-Ray Protective Shield:

Req. Quantity 1 unit

Description:

- **The shield should be compatible with BRANSIST ALEXA (Type-F12) -SHIMADZU-Digital angiography system's imaging positions.**
- The flange of the shield should be mountable with CATHLAB ceiling z-channel (drawing attached).
- It should be hand free and movable in any direction.
- It should have a suspension arm to control the appropriate direction.
- Stains of blood, contrast, bile etc., should be easily washable.
- Lead equivalence should be below 1.0 mm (Pb) or 0.50
- Radiation protection apparel should incorporate a high-quality protection sheet of even thickness.
- It should consist of Lead-free protective material manufactured with nontoxic heavy metals.
- It should be compatible with all generations of people – universal size.

Standards and Requirements:

- USFDA, CE and or ISO certifications or relevant standards certification.
- European CE Certified under PPE (Personal Protection Equipment) directive 89/686/EEC. The valid certificate from Notified body of EU with the Notified body no. on the certificate shall be enclosed.
- The supplier should visit the site and install the monitor.
- User/Application training to end users and In-house engineers should be provided by authorized trainers after the installation.

Additional Remarks or Requirements:

- Complete product details to be enclosed with the original brochure or catalogue (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, consumables and minimum supplies to be stated clearly.



3. X-Ray Protective Apron:

Required Numbers:

Quantity

- | | |
|---------------------------|----|
| • Lead Free Apron Size-XL | 05 |
| • Lead Free Apron Size-L | 05 |
| • Lead Free Apron Size-M | 05 |

Description:

- Apron should be Ultralight in weight.
- All models should have anti-skid pads resulting in comfort of wearing over shoulders.
- Should be frontal overlap.
- Adjustable in-built elastic belt should be there for reducing back and shoulder stress.
- Should be a slide slit for better mobility
- Should be an anti-skid shoulder pads for added comfort and weight distribution.
- Radiation protection apparel should incorporate high quality protection sheet of even thickness.
- Stains of blood, contrast, bile etc. should be easily washable.
- Apron should consist of Lead-free protective material manufactured with nontoxic heavy metals.
- Core material should be mix of Bilayer (Antimony, Bismuth, Tin, Barium, Titanium and Tungsten) having an Area Density of 2.8 Kg/sqm at 0.25 mm Pb.
- There should be absence of toxic material Lead and it should be Eco friendly.
- The supplier must provide color choice and department/Username on the aprons.
- Should be able to easy wear and remove.
- Should meet the Lead equivalence ranges 0.25mm Pb, 0.35mm Pb, 0.50 mm Pb adhesive Backing.
- Should be hook and loop type (Velcro).

Standards and Power Requirements:

- USFDA approved product. All other accessories should be USFDA, CE and or ISO certifications or relevant standards certification.
- European CE Certified under PPE (Personal Protection Equipment) directive 89/686/EEC. The valid certificate from Notified body of EU with the Notified body no. on the certificate shall be enclosed.

Additional Remarks or Requirements:

- Complete product details to be enclosed with the original brochure or catalogue. (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, consumables and minimum supplies to be stated clearly.



4. Lead Free Thyroid shield:

Req. Quantity **8 nos**

Description:

- Thyroid shield should be Ultralight in weight.
- Adjustable in-built elastic belt should be there for reducing back and shoulder stress.
- Should be a slide slit for better mobility
- Should be an anti-skid shoulder pads for added comfort and weight distribution.
- Radiation protection apparel should incorporate high quality protection sheet of even thickness.
- Stains of blood, contrast, bile etc. should be easily washable.
- Apron should consist of Lead-free protective material manufactured with nontoxic heavy metals.
- Core material should be mix of Bilayer (Antimony, Bismuth, Tin, Barium, Titanium and Tungsten) having an Area Density of 2.8 Kg/sqm at 0.25 mm Pb.
- There should be absence of toxic material Lead and it should be Eco friendly.
- The supplier must provide color choice and department/Username on the aprons.
- Should be able to easy wear and remove.
- Should meet the Lead equivalence ranges 0.25mm Pb, 0.35mm Pb, 0.50 mm Pb adhesive Backing.
- Should be hook and loop type (Velcro).
- Thyroid guard should be elegant or slimline.
- Should be easily attachable to all aprons

Standards and Power Requirements:

- USFDA approved product. All other accessories should be USFDA, CE and or ISO certifications or relevant standards certification.
- European CE Certified under PPE (Personal Protection Equipment) directive 89/686/EEC. The valid certificate from Notified body of EU with the Notified body no. on the certificate shall be enclosed.

Additional Remarks or Requirements:

- Complete product details to be enclosed with the original brochure or catalogue. (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, consumables and minimum supplies to be stated clearly.



5. Eye Wear/ Lead Goggle:

Req. Quantity **8 nos**

Description:

- Should be able to wear over regular spectacles.
- Lead equivalence should be 0.75 mm (Pb).
- It should be light in weight and comes in a size that fits all faces and shapes.
- Should have padding on the temples and bridge of the nose for comfort
- It should be available in different sizes for both adults and children.
- Stains of blood, contrast, bile etc. should be easily washable
- The supplier must provide colour choice.
- It should be lightweight

Standards and Power Requirements:

- USFDA approved product. All other accessories should be USFDA, CE and or ISO certifications or relevant standards certification.
- European CE Certified under PPE (Personal Protection Equipment) directive 89/686/EEC. The valid certificate from Notified body of EU with the Notified body no. on the certificate shall be enclosed.

Additional Remarks or Requirements:

- Complete product details to be enclosed with the original brochure or catalogue. (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, consumables and minimum supplies to be stated clearly.



6. X-Ray Protective Apron Storage:

Req. Quantity **3 nos**

Description:

- It should be mobile with a caster wheel and sturdy.
- It should accommodate at least 10 Lead Aprons
- Hangers' material should be stainless steel or stain-free and have adequate load-bearing capacity.
- It should be convenient to move with the braking mechanism.
- It should be easily accessible.
- Should be available detachable hooks for more storage options.
- Should be available the compact unit for storing apparel, gloves, and shields.

Standards and Power Requirements:

- USFDA approved product. All other accessories should be USFDA, CE and or ISO certifications or relevant standards certification.
- European CE Certified under PPE (Personal Protection Equipment) directive 89/686/EEC. The valid certificate from Notified body of EU with the Notified body no. on the certificate shall be enclosed.

Additional Remarks or Requirements:

- Complete product details to be enclosed with the original brochure or catalogue. (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, consumables and minimum supplies to be stated clearly.



7. UPS with Battery:

Req. Quantity 1 nos

Description:

- It should be suitable with the already purchased machine with at least 30 min battery backup.
- The battery should be available along with the UPS and match the number of batteries for the machine's electrical requirements.
- The minimum requirement should be 120kv or above as per the power requirements of using Cath-lab.
- The UPS panel should be a self-supporting stand-alone panel with a mounting facility to install over the floor.
- Should follow the requirements of output and input voltage as mentioned below.
 - Input
 - A. Voltage: - 3 phase 415 V AC Supply
 - B. Frequency: 50 Hz
 - Output
 - Voltage: - 240 V AC + 5 %
 - Frequency: - 50 Hz + 3 %
 - Power: - As per the machine requirements**
 - KW rating - more than 0.7 times of respective 120 KW**
 - Phase: - 1 phase
 - Total Harmonic Distortion < 3 %
 - Individual Harmonic Distortion < 5 %
 - Line regulation < 5 %
 - Load regulation < 5 %
 - Startup time < 1 milliseconds
- The environment range should meet the following requirements.: -
 - A. Temperature: - 0 to 450 centigrade
 - B. Humidity: - 90 % relative
 - C. Cooling: - forced cooling
- Efficiency should be more than 90 %
- The controlling unit should be Solid state Control.
- Should have AC-DC converter – Fully controlled bridge rectifier
- Should be DC-AC conversion – MOSFET / IGBT-based inverter (Party must specify the technique)
- Ripple in the regulator circuit output should be 4 V AC Peak to Peak maximum.
- DC battery charging voltage should be user selectable by utilizing a potentiometer on the card (not on the front panel).
- All the indicators are of solid-state LED type; lamp type indicators are not acceptable.



- Should available the following LED indicators: -
 - A) AC mains – set of 3 LED for each Phase
 - B) Low output voltage
 - C) Low DC voltage
 - D) Rectifier failure
 - E) Inverter failure
 - F) Battery overcharge
 - G) Battery fully discharge
 - H) Boost charging
 - I) Float charging
 - J) Extended charging
 - K) Bypass mode on
- Protections should follow the below-mentioned conditions.
 - A) Low output voltage
 - B) Low DC voltage
 - C) Rectifier failure
 - D) Inverter failure
 - E) Battery current limit
 - F) Incorrect phase sequence (rectifier shall trip)
- Meters should have the following parameters
 - A) Input AC voltage
 - B) Output AC Voltage
 - C) Output AC
 - D) Battery/DC voltage
 - E) Battery / DC current
- Controls switch should be available such as: -
 - A) AC mains ON / OFF
 - B) Inverter ON/ OFF
 - C) Battery Connect/ Disconnect
 - D) Load ON/OFF
 - E) Indicator test
 - F) Alarm acknowledgement
 - G) Alarm Reset
 - H) Battery charging auto /manual /boost mode
- A cooling fan should be available for power circuits.
- An exhaust fan should be available for cooling the overall equipment.
- Auto-detection should be available of battery charging option float /boost/ extended.



Standards and Power Requirements:

- USFDA, CE and or ISO certifications Security (EN50091-1) / EMC (EN50091-2) / IP Class (IP 20) / VFI Class (IEC62040-3) or relevant standards certification.
- Complete product details to be enclosed with the original brochure or catalogue (Soft or hard copy).
- Details of the standard accessories, additional accessories, optional items, Consumables and minimum supplies are to be stated clearly.