



EU DECLARATION OF CONFORMITY

Distributed by: JBM CAMPLLONG, S.L.U.

Address: CIM La Selva - Crta. Aeropuerto Km 1.6 Nave 2.2, 17185 Vilobí d'Onyar, Girona

VAT ID number: B17419292

Product Description: BATTERY TESTER

Manufacturer's part number: BT-005

Distributor reference: 51235

The submitted products have been tested by us with the following standard(s) and found to be in compliance with the listed European Directives:

Norma	Título	Edición/Fecha
EN IEC 61326-1	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements (Endorsed by Asociación Española de Normalización in July of 2021.)	2021
EN IEC 61000-3-2	Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions (equipment input current =16 A per phase)	2019+A1:2021
EN 61000-3-3	Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current =16 A per phase and not subject to conditional connection	2013+A1:2019+A2:2021
EN 61000-4-2	Electromagnetic compatibility (EMC) -- Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	2009
EN IEC 61000-4-3	Electromagnetic compatibility (EMC) - Part 4-3 : Testing and measurement techniques - Radiated, radio-frequency, electromagnetic field immunity test (Endorsed by Asociación Española de Normalización in April of 2021.)	2020
EN 61000-4-4	Electromagnetic compatibility (EMC) - Part 4-4: Testing and measurement techniques - Electrical fast transient/burst immunity test	2012
EN 61000-4-5	Electromagnetic compatibility (EMC) - Part 4-5: Testing and measurement techniques - Surge immunity test	2014+A1 : 2017
EN IEC 61000-4-6		2012
EN IEC 61000-4-8	Electromagnetic compatibility (EMC) -- Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity test	2010
EN IEC 61000-4-11	Electromagnetic compatibility (EMC) - Part 4-11: Testing and measurement techniques - Voltage dips, short interruptions and voltage variations immunity tests for equipment with input current up to 16 A per phase	2020

Signed by:



JBM CAMPLLONG, S.L.U.

CIM LA SELVA - Ctra. de l'Aeroport Km. 1,6 Nau 2.2

17185 VILLOBI D'ONYAR

GIRONA - SPAIN

Tel. 972 24 55 33 - Fax +34 972 24 54 37

Eduard Godoy

Purchasing Department Manager

Girona, December 19, 2023



Certificate of Conformity

Certificate Number: DL-20231016028C

Applicant:

Manufacturer:

Product: BATTERY TESTER

Model No.: BT-004
BT-001, BT-002, BT-003, BT-005, BT-006, BT-007T, BT-008, BT-009, BT-50

Test Standard: EN IEC 61326-1:2021
EN IEC 61000-3-2:2019+A1:2021, EN 61000-3-3:2013+A1:2019+A2:2021
EN 61000-4-2:2009, EN IEC 61000-4-3:2020, EN 61000-4-4:2012,
EN 61000-4-5:2014+A1:2017, EN IEC 61000-4-6:2022,
EN 61000-4-8:2010, EN IEC 61000-4-11:2020

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive. It is only valid in connection with the test report number: DL-20231016028E.

CE



This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole product and relevant. Without the written approval, It is not permitted to use the test lab's logo.

Shenzhen DL Testing Technology Co., Ltd.
101-201, Building C, Shuanghuan, No.8, Baoqing Road, Baolong Industrial Zone, Baolong Street,
Longgang District, Shenzhen, Guangdong, China

Web: www.dl-cert.com E-mail: Service@dl-cert.com Tel: 400-688-3552

