



DECLARACIÓN DE CONFORMIDAD

Distribuidor: JBM CAMPLLONG, S.L.U.

Dirección: CIM La Selva – Crta. Aeroport Km 1.6 Nave 2.2, 17185 Vilobí d'Onyar, Girona

CIF: B17419292

Descripción del producto: KIT DE PRUEBA DE ANTIGENO COVID-19

Referencia del fabricante: SL030101NST-1

Referencia JBM: 54118E

- El objeto de la declaración es conforme a la Directiva 98/79/EC sobre productos sanitarios para diagnóstico in vitro

Firmado:



Eduard Godoy

Director departamento de compras

En Girona, a 14 de enero de 2022



CERTIFICATE

EC Certificate No. 1434-IVDD-445/2021

**EC Design-examination
Directive 98/79/EC concerning
in vitro diagnostic medical devices**

Polish Centre for Testing and Certification certifies
that manufactured by:

Anhui Deepblue Medical Technology Co., Ltd.
4th Floor,D-1#Zone, Pearl Industrial Park, 106
Innovation Avenue, High-Tech Development Zone,
230088 Hefei, Anhui, China

in vitro diagnostic medical devices
for self-testing

COVID-19 (SARS-COV-2) Antigen Test Kit (Colloidal Gold)

SL030101NST-1,SL030101NST-2, SL030101NST-3, SL030101NST-5, SL030101NST-6, SL030101NST-7, SL030101NST-8,
SL030101NST-9, SL030101NST-10, SL030101NST-11, SL030101NST-12, SL030101NST-15, SL030101NST-16, SL030101NST-
17, SL030101NST-18, SL030101NST-19, SL030101NST-20, SL030101NST-25

in terms of design documentation, comply with requirements
of Annex III (Section 6) to Directive 98/79/EC (as amended)
implemented into Polish law,
as evidenced by the audit conducted by the PCBC

Validity of the Certificate: from 30.07.2021 to 27.05.2024

The date of issue of the Certificate: 30.07.2021

The date of the first issue of the Certificate: 22.07.2021



Issued under the Contract No. MD-96/2021
Application No: 183a/2021
Certificate bears the qualified signature.
Warsaw, 30.07.2021
Module A1

Anna
Małgorzata
Wyroba
Vice-President
Elektronicznie
podpisany przez Anna
Małgorzata Wyroba
Data: 2021.07.30
10:31:11 +02'00'



www.dbluemedical.com

DECLARATION OF CONFORMITY

MANUFACTURER: ANHUI DEEPBLUE MEDICAL TECHNOLOGY CO.,LTD.
4th Floor,D-1# Zone, Pearl Industrial Park, 106 Innovation Avenue,
High-Tech Development Zone , 230088 Hefei, Anhui, People's
Republic of China

EUROPEAN REPRESENTATIVE: Luxus Lebenswelt GmbH
Kochstr. 1, 47877, Willich, Germany

PRODUCT: COVID-19 (SARS-CoV-2) Antigen Test Kit (Colloidal Gold)

Models: SEE ATTACHMENT

REF: SEE ATTACHMENT

CLASSIFICATION: SELF-TESTING

EDMA CODE: 15 70 90 90 00

CONFORMITY ASSESSMENT ROUTE: Following the procedure relating to the EC Declaration of Conformity set out in Annex III Article 6 of Directive 98/79/EC.

WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCTS MEET THE PROVISIONS OF THE COUNCIL DIRECTIVE 98/79/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.

THE MANUFACTURER IS EXCLUSIVELY RESPONSIBLE FOR THE DECLARATION OF CONFORMITY.

STANDARDS APPLIED: EN ISO 13485:2016
EN ISO 18113-1:2011, EN ISO 18113-4:2011, EN 13612:
2002/AC:2002, EN ISO 23640:2015, EN 13641: 2002, EN ISO
15223-1: 2016, EN 13975:2003, EN 13532:2002, EN ISO
14971:2012.


NOTIFIED BODY: Polish Center for Testing and Certification
469 Puławska Street,02-844 Warsaw,Poland

(EN) CERTIFICATE(S): 1434-IVDD-445/2021

START OF CE-MARKING: 2021-07-30

PLACE, DATE OF ISSUE: HEFEI, 2021-10-27

SIGNATURE: CHEN FENGLING
GENERAL MANAGER



EC Declaration of Conformity


DOC-COVID-19 Ag(P/1)



www.dbluemedical.com

DECLARATION OF CONFORMITY ATTACHMENT

Specification	REF
1 piece per box	SL030101NST-1
2 pieces per box	SL030101NST-2
3 pieces per box	SL030101NST-3
5 pieces per box	SL030101NST-5
6 pieces per box	SL030101NST-6
7 pieces per box	SL030101NST-7
8 pieces per box	SL030101NST-8
9 pieces per box	SL030101NST-9
10 pieces per box	SL030101NST-10
11 pieces per box	SL030101NST-11
12 pieces per box	SL030101NST-12
15 pieces per box	SL030101NST-15
16 pieces per box	SL030101NST-16
17 pieces per box	SL030101NST-17
18 pieces per box	SL030101NST-18
19 pieces per box	SL030101NST-19
20 pieces per box	SL030101NST-20
25 pieces per box	SL030101NST-25





Certificate

No. Q5 003706 0001 Rev. 01

Holder of Certificate: **ANHUI DEEPBLUE MEDICAL TECHNOLOGY CO.,LTD.**
4th Floor, D-1# Zone
Pearl Industrial Park
106 Innovation Avenue, High-Tech Development Zone
230088 Hefei, Anhui
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate: **Design and Development, Production and Distribution of In Vitro Diagnostic Reagents by Colloidal Gold and Enzyme Chemical Reaction Method, Medical Ultrasonic Couplant, Acetowhite Solution, Epithelial Tissue Staining Solution, Rapid Test for Vaginitis(Polyamines) and Cell Preservation Solution**

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:Q5 003706 0001 Rev. 01

Report No.: SH21130301

Valid from: 2021-06-22

Valid until: 2024-06-21

Date, 2021-06-16

Christoph Dicks
Head of Certification/Notified Body

Certificate

No. Q5 003706 0001 Rev. 01

Applied Standard(s): EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

Facility(ies): ANHUI DEEPBLUE MEDICAL TECHNOLOGY CO.,LTD.
4th Floor, D-1# Zone, Pearl Industrial Park, 106 Innovation
Avenue, High-Tech Development Zone, 230088 Hefei, Anhui,
PEOPLE'S REPUBLIC OF CHINA

See Scope of Certificate