


Analysis of volatile free fatty acids in water

BP21 (FFAP)

Column part number	054462		
Phase	BP21 (FFAP)	Detector	FID
Column	25 m x 0.22 mm x 0.25 µm	Injection mode	Split
Initial temperature	Isothermal at 140°C		


 For more information visit www.trajanscimed.com or contact techsupport@trajanscimed.com