


Parallel GC for Complete Refinery Gas Analysis



Application

Hydrocarbon Processing

Author

Chunxiao Wang
Agilent Technologies (Shanghai) Co. Ltd.
412 Ying Lun Road
Waigaoqiao Free Trade Zone
Shanghai 200131
China

Abstract

An Agilent 7890A gas chromatograph configured with three parallel channels with simultaneous operation provides a complete, high-resolution analysis for refinery gas in six minutes. The system uses an optimized combination of several packed columns and PLOT alumina columns to allow fast separation of light hydrocarbons and permanent gases with the same oven temperature program. A third channel with TCD with nitrogen (or argon) carrier gas improves the hydrogen sensitivity and linearity. This application also shows the excellent performance for natural gas analysis.

Introduction

Refinery gas is a mixture of various gas streams produced in refinery processes. It can be used as a fuel gas, a final product, or a feedstock for further processing. An exact and fast analysis of the components is essential for optimizing refinery processes and controlling product quality. Refinery gas stream composition is very complex, typically containing hydrocarbons, permanent gases, sulfur compounds, and so on. Successful separation of such a complex gas mixture is often difficult using a single-channel GC system. Three parallel channel

analyses allow a separation problem to be divided into three sections. Each channel can optimize a particular part of the separation. TCD with helium carrier gas can be used for permanent gases analysis like O₂, N₂, CO, CO₂, H₂S, and COS. However, hydrogen has only a small difference in thermal conductivity compared to helium, making analysis by TCD using helium carrier gas difficult. To achieve full-range capability for hydrogen, an additional TCD with nitrogen or argon as a carrier is required. Light hydrocarbons are separated on an alumina PLOT column and detected on a FID.

The Agilent 7890A GC now supports an optional third detector (TCD), allowing simultaneous detection across three channels; this provides a complete analysis of permanent gases, including nitrogen, hydrogen, helium, oxygen, carbon monoxide, carbon dioxide, and hydrocarbons to nC₅, C₆+ fraction within six minutes.

Experimental

A single Agilent 7890A GC is configured with three channels, including one FID, and two TCDs. Light hydrocarbons are determined on the FID channel. One TCD with nitrogen or argon carrier is used for the determination of hydrogen and helium. The other TCD with helium carrier is used for the detection of all other required permanent gases. Figure 1 shows the valve drawing. The system conforms to published methods such as ASTM D1945 [1], D1946 [2], and UOP 539 [3].

The FID channel is for light hydrocarbon analysis. The sample from valve 4 is injected via the capillary injector into valve 3 to permit an early back-



flush of the grouped heavier hydrocarbons (normally C₆+). Valve 3 is a sequence reversal with a short DB1 (column 6) for separating the hexane plus fraction (C₆+) from the lighter components. C₁ through C₅ hydrocarbons are separated on a PLOT alumina column. As soon as the light components C₁ through C₅ pass through the DB1 column, valve 3 is switched to reverse the sequence of the DB1 and PLOT aluminum column so that components heavier than nC₆, including nC₆, are backflushed early. As a result, group C₆+ is followed by the individual hydrocarbons from the PLOT alumina column.


A new tube connector based on capillary flow technology is used to connect the valve to the capillary column to enhance the hydrocarbons analysis by improving the peak shape.

The second TCD channel (B TCD) employs three packed columns and two valves for the separation of permanent gases including O₂, N₂, CO, and CO₂ using helium as a carrier gas. Valve 1 is a 10-port valve used for gas sampling and backflushing heavier components; normally components heavier than ethylene are backflushed to vent when H₂S is not required to be analyzed. A six-port isolation

valve (valve 2) with adjustable restrictor is used to switch the molecular sieve 5A column in and out of the carrier stream. Initially, the isolated valve is in the OFF position so that unresolved components air, CO, and CH₄ pass quickly through the HayeSep Q (column 2) onto the molecular sieve (column 3). The valve is then switched to the ON position to trap them in column 3 and allow the CO₂ to bypass this column. When the CO₂ has eluted, valve 2 is switched back into the flow path to allow O₂, N₂, CH₄, and CO to elute from the molecular sieve column.

The third TCD channel (C TCD) is for the analysis of H₂. Sample from the 10-port valve (valve 5) is injected into a precolumn (column 4, HayeSep Q) when H₂ with its coeluted compounds O₂, N₂, and CO pass through the short precolumn HayeSep Q onto the molecular sieve 5A column (column 5). Valve 5 is switched so that CO₂ and other compounds will be backflushed to vent, while H₂ is separated on the molecular sieve 5A.

Typical GC conditions for fast refinery gas analysis are listed in Table 1. The refinery gas standard mixture that was used for the method development is listed in Table 2.



Column 1 HayeSep Q 80/100 mesh
 Column 2 HayeSep Q 80/100 mesh
 Column 3 Molsieve 5A 60/80 mesh
 Column 4 HayeSep Q 80/100 mesh

Column 5 Molsieve 5A 60/80 mesh
 Column 6 DB-1
 Column 7 HP-PLOT Al₂O₃
 PCM: Electronic pneumatics control (EPC) module

Figure 1. RGA valve system.

Table 1. Typical GC Conditions for Fast Refinery Gas Analysis

Valve temperature	120 °C
Oven temperature program	60 °C hold 1 min, to 80 °C at 20°C/min, to 190 °C at 30 °C/min
FID channel	
Front inlet	150°C, split ratio: 30:1 (uses higher or lower split ratio according to the concentrations of hydrocarbons)
Column	6: DB-1 7: HP-PLOT AI203 S
Column flow (He)	3.3 mL/min (12.7 psi at 60 °C), constant flow mode
FID	
Temperature	200 °C
H ₂ flow	40 mL/min
Air flow	400 mL/min
Make up (N ₂)	40 mL/min
Second TCD channel	
Column	1: HayeSep Q 80/100 mesh 2: HayeSep Q, 80/100 mesh 3: Molecular sieve 5A, 60/80 mesh
Column flow (He)	25 mL/min (36 psi at 60 °C), constant flow mode
Procolumn flow (He)	22 mL/min at 60 °C (7 psi), constant pressure mode
TCD	
Temperature	200 °C
Reference flow	45 mL/min
Make up	2 mL/min
Third TCD channel	
Column	4: HayeSep Q 80/100, mesh 5: Molecular sieve 5A, 60/80, mesh
Column flow (N ₂)	24 mL/min, (26 psi at 60 °C), constant flow mode
Procolumn flow (N ₂)	7 psi, (24 mL/min at 60 °C), constant pressure mode
TCD	
Temperature	200 °C
Reference flow	30 mL/min
Make up	2 mL/min

Table 2. RGA Calibration Gas Standards

	Compound	% (V/V)	Compound	% (V/V)	
1	Methane	5.98	15	i-Pentane	0.101
2	Ethane	5.07	16	n-pentane	0.146
3	Ethylene	2.99	17	1,3-Butadiene	1.46
4	Propane	8.04	18	Propyne	0.476
5	Cyclopropane	0.50	19	t-2-Pentene	0.195
6	Propylene	3.04	20	2-Methyl-2-butene	0.149
7	i-Butane	2.71	21	1-Pentene	0.094
8	n-Butane	2.11	22	c-2-Pentene	0.146
9	Propadiene	0.94	23	n-Hexane	0.099
10	Acetylene	1.72	24	H ₂	15.00
11	t-2-Butene	1.55	25	O ₂	2.00
12	1-Butene	1.00	26	CO	1.50
13	i-Butene	0.808	27	CO ₂	3.00
14	c-2-Butene	1.230	28	N ₂	BL

Results and Discussion

Enhance Gas Analysis with Union Connector

The system uses the new union connector based on capillary flow technology for connecting the capillary column to the valve, enhancing the peak shapes in gas analysis and making the connections easier. Figure 2 shows the comparison of peak shapes obtained from a traditional polyamide connector and the new union connector. With the new union connector the improvement in peak shape is readily apparent.

Fast Refinery Gas Analysis (RGA)

Use of an optimized combination of several packed columns and a PLOT alumina column allows fast separation of light hydrocarbons and permanent gases with the same oven temperature program without the need of an additional oven.

The separation results from each channel are illustrated in Figure 3.




Figure 2. Hydrocarbon peaks obtained from traditional tube connector and new union connector.




Figure 3. Refinery gas calibration standards analysis. The concentrations for each compound are shown in Table 2.

The top chromatogram (FID channel) is the hydrocarbon analysis. The PLOT alumina column provides excellent separation of hydrocarbons from C₁ to nC₅, including 22 isomers. Components heavier than nC₆ are backflushed early as a group (C₆+) through the precolumn. The middle chromatogram (second TCD channel) is the separation of permanent gases using helium as a carrier gas. The bottom chromatogram (third TCD channel) is the

separation of hydrogen, since hydrogen has only a little difference in thermal conductivity compared to helium. Use of an additional TCD with nitrogen (or argon) as a carrier gas improves the hydrogen detectability and linearity.

Table 3 shows very good repeatability for both retention time and area for analysis of the refinery gas standard.

Table3. Repeatability-Refinery Gas Analysis (6 runs) with 1 Run Excluded

Compounds	Retention time			Average	Area	
	Average	Std. dev.	RSD%		Std. dev.	RSD%
C ₆ +	0.99648	0.00031	0.03	59.01	1.10	1.86
Methane	1.50780	0.00046	0.03	490.02	1.45	0.30
Ethane	1.70788	0.00052	0.03	807.40	2.35	0.29
Ethylene	1.95732	0.00071	0.04	472.31	1.31	0.28
Propane	2.41706	0.00075	0.03	1950.35	5.96	0.31
Cyclopropane	3.18506	0.00075	0.02	145.62	0.45	0.31
Propylene	3.26195	0.00072	0.02	732.90	2.01	0.27
i-butane	3.64883	0.00055	0.02	885.04	3.15	0.36
n-butane	3.79161	0.00070	0.02	682.13	2.59	0.38
Propadiene	3.86098	0.00095	0.02	109.08	0.65	0.60
Acetylene	3.96990	0.00120	0.03	348.17	2.39	0.69
t-2-butene	4.47301	0.00106	0.02	507.88	2.59	0.51
1-butene	4.57118	0.00110	0.02	332.39	2.03	0.61
i-butylene	4.67529	0.00121	0.03	260.95	1.95	0.75
c-2-butene	4.76367	0.00112	0.02	403.80	3.47	0.86
i-pentane	5.03923	0.00090	0.02	45.03	0.05	0.11
n-pentane	5.14583	0.00099	0.02	69.23	0.40	0.58
1,3-butadiene	5.25906	0.00122	0.02	485.49	3.66	0.75
Propyne	5.36385	0.00155	0.03	101.08	0.41	0.40
t-2-pentene	5.58664	0.00121	0.02	82.85	0.66	0.79
2-methyl-2-butene	5.68220	0.00117	0.02	62.54	0.61	0.98
1-pentene	5.75553	0.00126	0.02	39.57	0.38	0.96
c-2-pentene	5.83970	0.00131	0.02	59.08	0.50	0.85
CO ₂	2.18561	0.00221	0.10	2040.33	2.37	0.12
O ₂	2.72634	0.00060	0.02	930.68	6.53	0.70
N ₂	3.25170	0.00044	0.01	22500.18	68.87	0.31
CO	4.61692	0.00083	0.02	903.09	2.77	0.31
H ₂	0.9869	0.00099	0.10	16097.38	106.53	0.66

Typical natural gas also can be characterized with the system using the same conditions for the fast RGA. The chromatograms of natural gas on the three channels are shown in Figure 4; hydrogen (3% Mol) and helium (1% Mol) are separated on the third TCD channel.

Flexibility for Hydrocarbon Analysis

The system is very flexible for hydrocarbon analysis. By setting up different valve (valve 3) switch times, the early backflush group can be C₆+ followed by individual C₁ to C₅ hydrocarbons as mentioned in fast RGA, or C₇+ followed by individual C₁ to C₆ hydrocarbons, or no backflush to separate C₁ to C₉ individual hydrocarbons. The top chromatogram in Figure 5 is the result with backflush group of C₆+, the middle one is that of C₇+, and the

bottom one is that of no backflush. With such flexibility, a wide range of refinery gas and natural gas compositions can be measured reliably without hardware or column changes.

H₂S and COS Analysis

H₂S and COS (methyl-mercaptan) can be analyzed on the rear TCD channel by adding an additional delay to the backflush time (valve 1) to allow H₂S and COS to elute onto column 2 (HayeSep Q). The analysis time is extended an additional 3 to 4 minutes, and requires a sample containing no water. Figure 6 shows the chromatogram of H₂S at approximately 500 ppm and COS 300 ppm with 1 mL sample size. The Nickel tubing packed columns and Hastelloy-C valves can be chosen for high concentration of H₂S analysis to minimize corrosion.




Figure 4. Natural gas analysis of a calibration gas.




Figure 5. Chromatograms of light hydrocarbons on FID channel with different backflush times .




Figure 6. H₂S at approximately 500 ppm and COS 300 ppm on second TCD channel.

Oven program: 50 hold 2 minutes, to 150 °C at
30 °C/min, hold 3 minutes, to
190 °C at 30 °C/min, hold 1 minute
Sample loop: 1 mL

Reporting

A macro program provides automated gas properties calculation. It gives a report in mole %, weight %, volume %, or any combination of the three. If required, heat values for the gas analyzed and other standard calculations are also available. Reports can be calculated using formulas given in the ASTM/GPA or ISO standards.

Conclusions

An exact and fast analysis of the components in refinery gas is essential for optimizing refinery processes and controlling product quality.

One 7890A GC configured with three parallel channels with simultaneous operation provides complete analysis of permanent gases, including nitrogen, hydrogen, helium, oxygen, carbon monoxide, carbon dioxide, and all hydrocarbons to C₅ and C₆₊ as a group within six minutes. A second TCD with nitrogen or argon as a carrier gas improves the hydrogen sensitivity and linearity.

The configuration is very flexible for hydrocarbon analysis, different backflush times may be set to obtain the early backflush group for C₆₊ or C₇₊, or no backflush to separate C₁ to C₁₀ individual hydrocarbons. In these cases, the analysis time is increased by 6 minutes. H₂S and COS can be analyzed on the same GC configuration; it requires 3 to 4 minutes of additional time.

A macro program provides automated gas properties calculation. Reports can be calculated using formulas given in the ASTM/GPA or ISO standards. It gives a report in mole %, weight %, volume %, or any combination of the three.

References

1. ASTM D1945-03, "Standard Test Method for Analysis of Natural Gas by Gas Chromatography," ASTM International, 100 Bar Harbor Drive, West Conshohocken, PA 19428 USA.
2. ASTM D1946-90 (2006), "Standard Practice for Analysis of Reformed Gas by Gas Chromatography," ASTM International, 100 Bar Harbor Drive, West Conshohocken, PA 19428 USA.
3. UOP Method 539, "Refinery Gas Analysis by Gas Chromatography," ASTM International, 100 Bar Harbor Drive, West Conshohocken, PA 19428, USA.

For More Information

For more information on our products and services, visit our Web site at www.agilent.com/chem.

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc. 2007

Printed in the USA
September 26, 2007
5989-7437EN

