


medis

MABS110BU SLIMMING MASSAGER

- Die Schichten zu lösen
- Sich von Ermüdung
- Nervenschmerzen
- Und Linderung der peripheren Nerven zu erzielen
- Den Blutkreislauf zu verbessern
- Den Muskeldruckern anmäßig zu lösen



Wichtige wichtige Informationen:
Nicht in folgenden Situationen benutzen:
Personen, die an einer der folgenden Krankheiten leiden oder irgendwelche Symptome dafür aufweisen, sollten das elektronische Impulsstromgerät für die Niederfrequenzbehandlung nicht benutzen.

1. Herzkrankheiten
2. Träger von Herzschrittmachern
3. Hohes Fieber
4. Akute (schmerzhaft) Erkrankungen
5. Unnormaler Blutdruck
6. Menstruation, Schwangerschaft oder nach der Geburt
7. An Verletzungen und/oder Hauterkrankungen oder um sie herum
8. Bösartige Tumore
9. Wenn der Körper vom Baden/Duschen oder Schwitzen nasst ist
10. Tuberkuloseerkrankungen
11. Leiste, die mit Medikamenten behandelt werden, sollten mit dem Arzt sprechen, bevor Sie das Massagegerät benutzen

Features:
Output of channels "A&B" can be used to treat various parts of the body simultaneously. You can choose any of the 5 treatment modes (tapping, modulation, kneading, program hand programs) or an even more comfortable massage.

Caution:
Do not use the unit in the following situations. These people who suffer from any of the following diseases or are showing any symptoms should not use electronic pulse massager for low-frequency treatment.

1. Heart diseases
2. Users of pacemakers
3. High fever
4. Acute (painful) diseases
5. Abnormal blood pressure
6. Menstruation, pregnancy or after giving birth
7. On or around cuts and/or skin diseases
8. Malignant tumors
9. When the body is wet from bathing/showering or sweating
10. Tuberculosis diseases
11. People receiving medical treatment should consult with physician prior to using the massager

Caution:
1. Do not use the massager at the same time, please ensure that each user uses only pads connected to one channel. It is advisable to always start with a low intensity setting with gradual increases to prevent any discomfort.

2. After the pads are on they do not overlap one another.
3. Never so turn to move the electrode pads to another part of your body or to another person's body without first stopping the power.
4. Always keep the electrode pads dry, do not touch the adhesive surface with your fingertips.

Body:
① Electrode jack
② Intensity scale
③ Mode selector
④ Speed selector
⑤ Open/refresh key
⑥ On/Off key
⑦ Mode key
⑧ Intensity/Speed key

Correct use:
1. When replacing the batteries, be sure that the power is off.

2. Open the battery compartment by pushing it gently in the direction of the arrows.

3. Insert the battery pack into the battery compartment. Refer to the diagram indicated at the back of the unit. Slide the battery cover completely shut.

4. Insert AA (AAA) V.

Preparing & setting the electrode cord:
1. Carefully connect the electrode cord to the electrode pads.

2. Insert the plug of the electrode cord into the unit's electrode jack.

3. Peel the film from the electrode pads.

4. Use clean towels to wipe the area of skin which you will be affixing the pads to remove any oil, cosmetic or dirt.

5. Apply the electrode pads to select both their adhesiveness, and the number of times you wish to use the reduced.

6. Show the electrode pads to the area of the body you wish to massage. The unit cannot operate unless both pads from at least one channel are applied.

Caution:
1. Always remove batteries if the unit is not in use for a long time.

2. Never apply the electrode pads to your skin with the power supply on. Doing so may result in sudden shock.

3. Do not use any areas of the body which the pads cannot be easily affixed.

4. For this use, surgical tape from your local pharmacy to secure the pads.

Usage:
1. Turn the power on by pressing the power switch. The TAP symbol will appear on the LCD screen.

2. Select the desired treatment mode by pressing the MODE button.

3. Raise the intensity gradually by twisting the intensity control for both channels A and B. Both channels can be set separately. Intensity will increase as you increase the intensity. When you change the treatment mode, the stimulation control for each channel will stop. Set the intensity for the new mode by twisting the intensity control again.

4. After treatment, turn off the power button.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Merkmale:

Der Ausgang von 2 Kanälen (A & B) kann zur gleichzeitigen Behandlung von Körperstellen benutzt werden, wobei dabei eine Qualität des Modalitäten (Klopfen, Modulation, Kneten, Programm 1 & Programm 2) ausgewählt. Die Wellenform und -frequenz wechelt automatisch zwischen 2 Computerprogrammen, um eine bequemer Massage zu gewährleisten.

Benutzen Sie das MABS110BU -Abnehmungsgerät, um:

- Die Schichten zu lösen
- Sich von Ermüdung
- Nervenschmerzen
- Und Linderung der peripheren Nerven zu erzielen
- Den Blutkreislauf zu verbessern
- Den Muskeldruckern anmäßig zu lösen

Wichtige wichtige Informationen:
Nicht in folgenden Situationen benutzen:
Personen, die an einer der folgenden Krankheiten leiden oder irgendwelche Symptome dafür aufweisen, sollten das elektronische Impulsstromgerät für die Niederfrequenzbehandlung nicht benutzen.

1. Herzkrankheiten
2. Träger von Herzschrittmachern
3. Hohes Fieber
4. Akute (schmerzhaft) Erkrankungen
5. Unnormaler Blutdruck
6. Menstruation, Schwangerschaft oder nach der Geburt
7. An Verletzungen und/oder Hauterkrankungen oder um sie herum
8. Bösartige Tumore
9. Wenn der Körper vom Baden/Duschen oder Schwitzen nasst ist
10. Tuberkuloseerkrankungen
11. Leiste, die mit Medikamenten behandelt werden, sollten mit dem Arzt sprechen, bevor Sie das Massagegerät benutzen

Features:
Output of channels "A&B" can be used to treat various parts of the body simultaneously. You can choose any of the 5 treatment modes (tapping, modulation, kneading, program hand programs) or an even more comfortable massage.

Caution:
Do not use the unit in the following situations. These people who suffer from any of the following diseases or are showing any symptoms should not use electronic pulse massager for low-frequency treatment.

1. Heart diseases
2. Users of pacemakers
3. High fever
4. Acute (painful) diseases
5. Abnormal blood pressure
6. Menstruation, pregnancy or after giving birth
7. On or around cuts and/or skin diseases
8. Malignant tumors
9. When the body is wet from bathing/showering or sweating
10. Tuberculosis diseases
11. People receiving medical treatment should consult with physician prior to using the massager

Caution:
1. Do not use the massager at the same time, please ensure that each user uses only pads connected to one channel. It is advisable to always start with a low intensity setting with gradual increases to prevent any discomfort.

2. After the pads are on they do not overlap one another.

3. Never so turn to move the electrode pads to another part of your body or to another person's body without first stopping the power.

4. Always keep the electrode pads dry, do not touch the adhesive surface with your fingertips.

5. Do not use any areas of the body which the pads cannot be easily affixed.

6. For this use, surgical tape from your local pharmacy to secure the pads.

Body:
① Electrode jack
② Intensity scale
③ Mode selector
④ Speed selector
⑤ Open/refresh key
⑥ On/Off key
⑦ Mode key
⑧ Intensity/Speed key

Correct use:
1. When replacing the batteries, be sure that the power is off.

2. Open the battery compartment by pushing it gently in the direction of the arrows.

3. Insert the battery pack into the battery compartment. Refer to the diagram indicated at the back of the unit. Slide the battery cover completely shut.

4. Insert AA (AAA) V.

Preparing & setting the electrode cord:
1. Carefully connect the electrode cord to the electrode pads.

2. Insert the plug of the electrode cord into the unit's electrode jack.

3. Peel the film from the electrode pads.

4. Use clean towels to wipe the area of skin which you will be affixing the pads to remove any oil, cosmetic or dirt.

5. Apply the electrode pads to select both their adhesiveness, and the number of times you wish to use the reduced.

6. Show the electrode pads to the area of the body you wish to massage. The unit cannot operate unless both pads from at least one channel are applied.

Caution:
1. Always remove batteries if the unit is not in use for a long time.

2. Never apply the electrode pads to your skin with the power supply on. Doing so may result in sudden shock.

3. Do not use any areas of the body which the pads cannot be easily affixed.

4. For this use, surgical tape from your local pharmacy to secure the pads.

Usage:
1. Turn the power on by pressing the power switch. The TAP symbol will appear on the LCD screen.

2. Select the desired treatment mode by pressing the MODE button.

3. Raise the intensity gradually by twisting the intensity control for both channels A and B. Both channels can be set separately. Intensity will increase as you increase the intensity. When you change the treatment mode, the stimulation control for each channel will stop. Set the intensity for the new mode by twisting the intensity control again.

4. After treatment, turn off the power button.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Caratteristiche:

Le 2 canali de sortie A & B peuvent être utilisés pour traiter des parties différentes du corps de façon simultanée. Vous pouvez choisir entre 5 modes de traitement (tapement, modulation, pétrissage, programme 1 et programme 2). Le format et la fréquence de l'onde fonctionnent automatiquement dans deux programmes informatiques pour un massage encore plus confortable.

Utilisez le massoir amincissant MABS110BU pour:

- Détruire les couches adipeuses
- La récupération
- Les maigrir
- La prévention des nerfs périphériques
- Stimuler la circulation sanguine
- Soulager graduellement les douleurs musculaires

Informations préliminaires importantes:
Ne pas utiliser ce dispositif dans les situations suivantes: Les personnes qui souffrent de maladies suivantes ou qui ont montrent des symptômes ne doivent pas utiliser cet appareil, surtout dans les cas suivants:
1. Maladies cardiaques
2. Utilisateurs de pacemakers
3. Fièvre aiguë
4. Maladies aiguës (douleurs)
5. Pression artérielle anormale
6. Règles, grossesse ou après un accouchement
7. A proximité de coupures et/ou en cas de maladies cutanées
8. Tumeurs malignes
9. Si le corps est en sueur ou mouillé après un bain/une douche
10. Tuberculose existante
11. Les personnes en cours de traitement médical doivent consulter un médecin avant d'utiliser le massoir

Features:
Output of channels "A&B" can be used to treat various parts of the body simultaneously. You can choose any of the 5 treatment modes (tapping, modulation, kneading, program hand programs) or an even more comfortable massage.

Caution:
Do not use the unit in the following situations. These people who suffer from any of the following diseases or are showing any symptoms should not use electronic pulse massager for low-frequency treatment.

1. Heart diseases
2. Users of pacemakers
3. High fever
4. Acute (painful) diseases
5. Abnormal blood pressure
6. Menstruation, pregnancy or after giving birth
7. On or around cuts and/or skin diseases
8. Malignant tumors
9. When the body is wet from bathing/showering or sweating
10. Tuberculosis diseases
11. People receiving medical treatment should consult with physician prior to using the massager

Caution:
1. Do not use the massager at the same time, please ensure that each user uses only pads connected to one channel. It is advisable to always start with a low intensity setting with gradual increases to prevent any discomfort.

2. After the pads are on they do not overlap one another.

3. Never so turn to move the electrode pads to another part of your body or to another person's body without first stopping the power.

4. Always keep the electrode pads dry, do not touch the adhesive surface with your fingertips.

5. Do not use any areas of the body which the pads cannot be easily affixed.

6. For this use, surgical tape from your local pharmacy to secure the pads.

Body:
① Electrode jack
② Intensity scale
③ Mode selector
④ Speed selector
⑤ Open/refresh key
⑥ On/Off key
⑦ Mode key
⑧ Intensity/Speed key

Correct use:
1. When replacing the batteries, be sure that the power is off.

2. Open the battery compartment by pushing it gently in the direction of the arrows.

3. Insert the battery pack into the battery compartment. Refer to the diagram indicated at the back of the unit. Slide the battery cover completely shut.

4. Insert AA (AAA) V.

Preparing & setting the electrode cord:
1. Carefully connect the electrode cord to the electrode pads.

2. Insert the plug of the electrode cord into the unit's electrode jack.

3. Peel the film from the electrode pads.

4. Use clean towels to wipe the area of skin which you will be affixing the pads to remove any oil, cosmetic or dirt.

5. Apply the electrode pads to select both their adhesiveness, and the number of times you wish to use the reduced.

6. Show the electrode pads to the area of the body you wish to massage. The unit cannot operate unless both pads from at least one channel are applied.

Caution:
1. Always remove batteries if the unit is not in use for a long time.

2. Never apply the electrode pads to your skin with the power supply on. Doing so may result in sudden shock.

3. Do not use any areas of the body which the pads cannot be easily affixed.

4. For this use, surgical tape from your local pharmacy to secure the pads.

Usage:
1. Turn the power on by pressing the power switch. The TAP symbol will appear on the LCD screen.

2. Select the desired treatment mode by pressing the MODE button.

3. Raise the intensity gradually by twisting the intensity control for both channels A and B. Both channels can be set separately. Intensity will increase as you increase the intensity. When you change the treatment mode, the stimulation control for each channel will stop. Set the intensity for the new mode by twisting the intensity control again.

4. After treatment, turn off the power button.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Maintenance and storage of electrode pads:
1. If the electrode pads become soaked the adhesive power may decrease and the skin might become irritated. If this should happen moisten the surface of the pads with water and wipe away the water. Use a clean towel to dry the pads.

2. Do not use or store the unit where are magnetic fields or electric wave (near TV sets or speakers).

3. Do not place the main unit in areas of high temperature, high humidity or under direct sunlight.

4. Store the unit where there is no moisture.

5. Keep out of reach from children.

6. Remove the batteries if the unit will not be used for an extended period of time.

Eigenschaften:

Zwei Kanäle (Ausgang A & B) können zur gleichzeitigen Behandlung von Körperstellen benutzt werden, wobei dabei eine Qualität des Modalitäten (Klopfen, Modulation, Kneten, Programm 1 & Programm 2) ausgewählt. Die Wellenform und -frequenz wechelt automatisch zwischen 2 Computerprogrammen, um eine bequemer Massage zu gewährleisten.

Benutzen Sie das MABS110BU -Abnehmungsgerät, um:

- Die Schichten zu lösen
- Sich von Ermüdung
- Nervenschmerzen
- Und Linderung der peripheren Nerven zu erzielen
- Den Blutkreislauf zu verbessern
- Den Muskeldruckern anmäßig zu lösen

Wichtige wichtige Informationen:
Nicht in folgenden Situationen benutzen:
Personen, die an einer der folgenden Krankheiten leiden oder irgendwelche Symptome dafür aufweisen, sollten das elektronische Impulsstromgerät für die Niederfrequenzbehandlung nicht benutzen.

1. Herzkrankheiten
2. Träger von Herzschrittmachern
3. Hohes Fieber
4. Akute (schmerzhaft) Erkrankungen
5. Unnormaler Blutdruck
6. Menstruation, Schwangerschaft oder nach der Geburt
7. An Verletzungen und/oder Hauterkrankungen oder um sie herum
8. Bösartige Tumore
9. Wenn der Körper vom Baden/Duschen oder Schwitzen nasst ist
10. Tuberkuloseerkrankungen
11. Leiste, die mit Medikamenten behandelt werden, sollten mit dem Arzt sprechen, bevor Sie das Massagegerät benutzen

Features:
Output of channels "A&B" can be used to treat various parts of the body simultaneously. You can choose any of the 5 treatment modes (tapping, modulation, kneading, program hand programs) or an even more comfortable massage.

Caution:
Do not use the unit in the following situations. These people who suffer from any of the following diseases or are showing any symptoms should not use electronic pulse massager for low-frequency treatment.

1. Heart diseases
2. Users of pacemakers
3. High fever
4. Acute (painful) diseases
5. Abnormal blood pressure
6. Menstruation, pregnancy or after giving birth
7. On or around cuts and/or skin diseases
8. Malignant tumors
9. When the body is wet from bathing/showering or sweating
10. Tuberculosis diseases
11. People receiving medical treatment should consult with physician prior to using the massager

Caution:
1. Do not use the massager at the same time, please ensure that each user uses only pads connected to one channel. It is advisable to always start with a low intensity setting with gradual increases to prevent any discomfort.

2. After the pads are on they do not overlap one another.

3. Never so turn to move the electrode pads to another part of your body or to another person's body without first stopping the power.

4. Always keep the electrode pads dry, do not touch the adhesive surface with your fingertips.

5. Do not use any areas of the body which the pads cannot be easily affixed.

