



307 Lewers St., Suite 707  
Honolulu, HI 96815  
1-888-387-6344



## Certificate of Calibration

Date: Aug 5, 2025

Cert No. [REDACTED]

Customer:

Honolulu HI 96819

MPC Control #:

Asset ID: N/A

Gage Type: ANALYTICAL BALANCE

Manufacturer: METTLER TOLEDO

Model Number: MS1003S

Size: 1 KG

Temp/RH: 20.0°C / 40.0%

Location: Calibration performed at Customer's facility

Work Order #:

Purchase Order #:

Serial Number:

Department: N/A

Performed By: GLENN HIGGINS

Received Condition: IN TOLERANCE

Returned Condition: IN TOLERANCE

Cal. Date: August 04, 2025

Cal. Interval: 12 MONTHS

Cal. Due Date: August 04, 2026

### Calibration Notes:

### Standards Used to Calibrate Equipment

I.D.	Description.	Model	Serial	Manufacturer	Cal. Due Date	Traceability #
[REDACTED]	WEIGHT SET, CLASS F1	OIML CLASS F1	[REDACTED]	CHANGZHOU FUYUE WEIGHT CO	May 31, 2026	[REDACTED]

### Procedures Used in this Event

Procedure Name	Description
MPC-WEI-001 Rev. 09	Weighing Instruments, General, Rev.09, Jul-25-2025

Calibrating Technician:

GLENN HIGGINS

QC Approval:

ROBERT MEANS

**STATEMENTS OF PASS OR FAIL CONFORMANCE:** The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006

**THE CALIBRATION REPORT STATUS:**

PASS: Term used when compliance statement is given and the measurement result is PASS.

PASS<sup>2</sup>: Term used when compliance statement is given and the measurement result is conditionally passed or PASS<sup>2</sup>.

FAIL: Term used when compliance statement is given and the measurement result is FAIL.

FAIL<sup>2</sup>: Term used when compliance statement is given and the measurement result is conditionally failed or FAIL<sup>2</sup>.

**REPORT OF VALUE:** Term used when reported measurement is not requiring compliance statement in report.

**ADJUSTED:** When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as effect.

**LIMITED:** When an instrument fails a calibration but still functions in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2 which for a normal distribution corresponds to a coverage probability of approximately 95% unless otherwise stated. The calibration report complies with ISO/IEC 17025:2017 ANSI/NCSL Z540.3-2006 and ANSI/NCSL Z540.1-1994. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Re-calibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified. This may not be reproduced in part or in whole without the prior written approval of the issuing MP Calibration Laboratory.