


Figure S16

A

parameter	value
1 Cytokine	IL -2R
2 Vendor	Invitrogen
3 Samples included in this agreement	14
4 Proportion of both readings in range	37.8
5 Limits of agreement	0.48 to 2.23
6 Constant variance p.value	0.587
7 Constant ratio p.value	0.048
8 Ratio is 1 p.value	0.778



B

parameter	value
1 Cytokine	IL-2R
2 Vendor	INV_MAG
3 Samples included in this agreement	30
4 Proportion of both readings in range	75.0
5 Limits of agreement	0.64 to 1.45
6 Constant variance p.value	0.662
7 Constant ratio p.value	0.238
8 Ratio is 1 p.value	0.313

