





Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Masks with Magnetic Clips ¹	
	
AirFit™ F20 / AirFit™ F20 for Her	
Mask System	
63400	AirFit F20 Mask System - SML
63400	AirFit F20 Mask System - SML
63401	AirFit F20 Mask System - MED
63402	AirFit F20 Mask System - LGE
63403	AirFit F20 for Her Mask System - SML
63403	AirFit F20 for Her Mask System - SML
63404	AirFit F20 for Her Mask System - MED

Masks without Magnetic Clips			
			
AirFit™ F10 / AirFit™ F10 for Her		Mirage Quattro™	
Mask System			
63101	AirFit F10 Mask System - SML	61200	Mirage Quattro Mask System - X-SML
63101	AirFit F10 Mask System - SML	61201	Mirage Quattro Mask System - SML
63102	AirFit F10 Mask System - MED	61202	Mirage Quattro Mask System - MED
63103	AirFit F10 Mask System - LGE	61203	Mirage Quattro Mask System - LGE
63139	AirFit F10 for Her Mask System - X-SML	61200	Mirage Quattro Mask System - X-SML
63140	AirFit F10 for Her Mask System - SML	61201	Mirage Quattro Mask System - SML
63141	AirFit F10 for Her Mask System - MED	61202	Mirage Quattro Mask System - MED

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit ResMed.com for more information regarding these masks.




Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Masks with Magnetic Clips ¹	
	
AirTouch™ F20 / AirTouch™ F20 for Her	
Mask System	
63000	AirTouch F20 Mask System - SML
63000	AirTouch F20 Mask System - SML
63001	AirTouch F20 Mask System - MED
63002	AirTouch F20 Mask System - LGE
63003	AirTouch F20 for Her Mask System - SML
63003	AirTouch F20 for Her Mask System - SML
63004	AirTouch F20 for Her Mask System - MED

Masks without Magnetic Clips			
			
AirFit™ F10 / AirFit™ F10 for Her		Mirage Quattro™	
Mask System			
63101	AirFit F10 Mask System - SML	61200	Mirage Quattro Mask System - X-SML
63101	AirFit F10 Mask System - SML	61201	Mirage Quattro Mask System - SML
63102	AirFit F10 Mask System - MED	61202	Mirage Quattro Mask System - MED
63103	AirFit F10 Mask System - LGE	61203	Mirage Quattro Mask System - LGE
63139	AirFit F10 for Her Mask System - X-SML	61200	Mirage Quattro Mask System - X-SML
63140	AirFit F10 for Her Mask System - SML	61201	Mirage Quattro Mask System - SML
63141	AirFit F10 for Her Mask System - MED	61202	Mirage Quattro Mask System - MED

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit ResMed.com for more information regarding these masks.





Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Mask with Magnetic Clips ¹	
	
AirFit™ F30	
Mask System	
64100	AirFit F30 Mask System - SML
64100	AirFit F30 Mask System - SML
64101	AirFit F30 Mask System - MED
64101	AirFit F30 Mask System - MED

Masks without Magnetic Clips			
			
AirFit™ F10		Mirage Quattro™	
Mask System			
63101	AirFit F10 Mask System - SML	61200	Mirage Quattro Mask System - X-SML
63101	AirFit F10 Mask System - SML	61201	Mirage Quattro Mask System - SML
63102	AirFit F10 Mask System - MED	61202	Mirage Quattro Mask System - MED
63103	AirFit F10 Mask System - LGE	61203	Mirage Quattro Mask System - LGE

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit ResMed.com for more information regarding these masks.






Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Mask with Magnetic Clips ¹	
	
AirFit™ F30i	
Mask System	
63330	AirFit F30i Mask System - S/SML
63330	AirFit F30i Mask System - S/SML
63331	AirFit F30i Mask System - S/STD
63332	AirFit F30i Mask System - M/STD
63334	AirFit F30i Mask System - M/LGE
63333	AirFit F30i Mask System - W/STD

Masks without Magnetic Clips			
			
AirFit™ F10		Mirage Quattro™	
Mask System			
63101	AirFit F10 Mask System - SML	61200	Mirage Quattro Mask System - X-SML
63101	AirFit F10 Mask System - SML	61201	Mirage Quattro Mask System - SML
63101	AirFit F10 Mask System - SML	61201	Mirage Quattro Mask System - SML
63102	AirFit F10 Mask System - MED	61202	Mirage Quattro Mask System - MED
63102	AirFit F10 Mask System - MED	61202	Mirage Quattro Mask System - MED
63103	AirFit F10 Mask System - LGE	61203	Mirage Quattro Mask System - LGE

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit ResMed.com for more information regarding these masks.





Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Nasal masks

Masks with Magnetic Clips ¹	
	
AirFit™ N20 / AirFit™ N20 for Her	
Mask System	
63503	AirFit N20 Mask System - SML
63501	AirFit N20 Mask System - MED
63502	AirFit N20 Mask System - LGE
63502	AirFit N20 Mask System - LGE
63500	AirFit N20 for Her Mask System

Masks without Magnetic Clips			
			
AirFit™ N30		AirFit™ N30i²	
Mask System			
64222	AirFit N30 Mask System - SML	63801	AirFit N30i Starter Pack - SML
64223	AirFit N30 Mask System - MED	63800	AirFit N30i Starter Pack - STD
64223	AirFit N30 Mask System - MED	63800	AirFit N30i Starter Pack - STD
64224	AirFit N30 Mask System - SML-WIDE	63800	AirFit N30i Starter Pack - STD
64222	AirFit N30 Mask System - SML	63801	AirFit N30i Starter Pack - SML

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit ResMed.com for more information regarding these masks.





Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Nasal masks

Masks with Magnetic Clips ¹	
	
AirTouch™ N20 / AirTouch™ N20 for Her	
Mask System	
63903	AirTouch N20 Mask System - SML
63901	AirTouch N20 Mask System - MED
63902	AirTouch N20 Mask System - LGE
63902	AirTouch N20 Mask System - LGE
63900	AirTouch N20 for Her Mask System

Masks without Magnetic Clips			
			
AirFit™ N30		AirFit™ N30i²	
Mask System			
64222	AirFit N30 Mask System - SML	63801	AirFit N30i Starter Pack - SML
64223	AirFit N30 Mask System - MED	63800	AirFit N30i Starter Pack - STD
64223	AirFit N30 Mask System - MED	63800	AirFit N30i Starter Pack - STD
64224	AirFit N30 Mask System - SML-WIDE	63800	AirFit N30i Starter Pack - STD
64222	AirFit N30 Mask System - SML	63801	AirFit N30i Starter Pack - SML

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit ResMed.com for more information regarding these masks.



This content is intended for health professionals only. Please refer to the user guide for relevant information related to any warnings and precautions to be considered before and during use of the product.

1. The mask contains magnets that may interfere with certain implants or medical devices. They are not to be used by patients where they, or anyone in close physical contact while using the mask, have active medical implants that interact with magnets and/or metallic implants/objects containing ferromagnetic material. Keep the mask magnets at a safe distance of at least 6 inches (150 mm) away from implants or medical devices that may be adversely affected by magnetic interference. Please refer to the User Guide for complete information, including magnet contraindications and warnings.
2. AirFit™ N30i Starter Packs (both Small and Standard) are supplied with Small, Medium, and Small-Wide cushion sizes.